

Table of contents

\USER

CRM AW

GOBrain

GOBrain_Aera_20ch

AAHead_Scout
t1_se_rs_sag_p2
t2_tse_tra_p2
t2_tse_dark-fluid_tra
ep2d_diff_mddw_12_p2
ep2d_tra_hemo

\\USER\CRM AW\GOBrain\GOBrain_Aera_20ch\AAHead_Scout

TA: 0:19 PM: REF Voxel size: 1.6×1.6×1.6 mmPAT: 3 Rel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L0.0 P20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	4.52 ms
TE	2.38 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HE1-4

Contrast - Common

TR	4.52 ms
TE	2.38 ms
Flip angle	8 deg

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
Base resolution	160
Phase resolution	100 %
Slice resolution	69 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Trajectory	Cartesian

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	24
Accel. factor 3D	1

Resolution - iPAT

Reference scan mode	Integrated
---------------------	------------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L0.0 P20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	4.52 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	L0.0 P20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off

System - Miscellaneous

Coil Select Mode	Off - AutoCoilSelect
------------------	----------------------

System - Adjustments

B0 Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - Tx/Rx

Frequency 1H	63.670263 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Flip angle	8 deg
Measurements	1
Time to center	8.2 s

Inline - Inline

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Inline - MapIt

Save original images	On
MapIt	None
Flip angle	8 deg
Measurements	1
Contrasts	1
TR	4.52 ms
TE	2.38 ms

Sequence - Part 1

Introduction	On
Dimension	3D
Asymmetric echo	Weak
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	540 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Normal
Excitation	Non-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
------	-----

\\USER\CRM AW\GOBrain\GOBrain_Aera_20chlt1_se_rs_sag_p2

TA: 1:11 PM: REF Voxel size: 0.9×0.9×5.0 mmPAT: 2 Rel. SNR: 1.00 : se_rs

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Routine

Slice group	1
Slices	27
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Basis
Phase oversampling	10 %
FoV read	230 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	560.0 ms
TE	12.0 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize, Elliptical filter
Coil elements	HE1-4

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	27
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
FoV read	230 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	560.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Contrast - Common

TR	560.0 ms
TE	12.0 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Fat suppr.	None
Water suppr.	None
Restore magn.	Off

Geometry - AutoAlign

Slice group	1
AutoAlign	Head > Basis
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Restore magn.	Off
Special sat.	None

Resolution - Common

FoV read	230 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	256
Phase resolution	80 %
Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	Off

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

Geometry - Tim CT

Tim CT mode	Off
Slices	27
Slice thickness	5.0 mm
Dist. factor	20 %
FoV read	230 mm
FoV phase	100.0 %

System - Miscellaneous

Positioning mode	REF
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Basis
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - Tx/Rx

Frequency 1H	63.670263 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	560.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	230 mm
FoV phase	100.0 %
Phase resolution	80 %
Trajectory	Cartesian

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	1
Flow comp.	Slice
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0 ms
Bandwidth	150 Hz/Px

Sequence - Part 2

Define	Turbo factor
Echo trains per slice	125
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Fast
Gradient mode	Fast
WARP	Off
Red. EC sensitivity	Off
Turbo factor	1

Sequence - Assistant

Mode	Off
Allowed delay	30 s

\\USER\CRM AW\GOBrain\GOBrain_Aera_20ch\t2_tse_tra_p2

TA: 0:56 PM: REF Voxel size: 0.9×0.9×5.0 mmPAT: 2 Rel. SNR: 1.00 : tse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	39
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Routine

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
TR	4700.0 ms
TE	111 ms
Averages	1
Concatenations	2
Filter	Prescan Normalize, Elliptical filter
Coil elements	HE1-4

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
TR	4700.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	2

Contrast - Common

TR	4700.0 ms
TE	111 ms
TD	0.0 ms
MTC	Off
Magn. preparation	None
Flip angle	180 deg
Fat suppr.	None
Water suppr.	None
Restore magn.	Off

Geometry - AutoAlign

Slice group	1
AutoAlign	Head > Brain
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Restore magn.	Off
Special sat.	None

Resolution - Common

FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
Base resolution	256
Phase resolution	98 %
Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	Off

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

Geometry - Tim CT

Tim CT mode	Off
Slices	25
Slice thickness	5.0 mm
Dist. factor	20 %
FoV read	220 mm
FoV phase	87.5 %

System - Miscellaneous

Positioning mode	REF
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	90.00 deg
R >> L	193 mm
A >> P	220 mm
F >> H	149 mm
Reset	Off

System - Tx/Rx

Frequency 1H	63.670263 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	4700.0 ms
Concatenations	2

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	220 mm
FoV phase	87.5 %
Phase resolution	98 %
Trajectory	Cartesian

Physio - PACE

Resp. control	Off
Concatenations	2

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	11.1 ms
Bandwidth	190 Hz/Px

Sequence - Part 2

Define	Turbo factor
Echo trains per slice	5
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Low SAR
Gradient mode	Normal
WARP	Off
Red. EC sensitivity	Off
Turbo factor	26

Sequence - Assistant

Mode	Off
Allowed delay	30 s

\\USER\CRM AW\GOBrain\GOBrain_Aera_20ch\t2_tse_dark-fluid_tra

TA: 2:01 PM: REF Voxel size: 0.4×0.4×5.0 mmPAT: Off Rel. SNR: 1.00 : tir

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
TR	5500.0 ms
TE	78.0 ms
Averages	1
Concatenations	2
Filter	Prescan Normalize, Elliptical filter
Coil elements	HE1-4

Contrast - Common

TR	5500.0 ms
TE	78.0 ms
TD	0.0 ms
MTC	Off
Magn. preparation	Slice-sel. IR
TI	1930 ms
Flip angle	180 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong
Water suppr.	None
Restore magn.	Off
Freeze suppressed tissue	On

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
Base resolution	256
Phase resolution	100 %

Resolution - Common

Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	On

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
TR	5500.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	2

Geometry - AutoAlign

Slice group	1
AutoAlign	Head > Brain
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Water suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H

Geometry - Tim Planning Suite

Table position	0 mm
Inline Composing	Off

Geometry - Tim CT

Tim CT mode	Off
Slices	25
Slice thickness	5.0 mm
Dist. factor	20 %
FoV read	220 mm
FoV phase	87.5 %

System - Miscellaneous

Positioning mode	REF
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	90.00 deg
R >> L	193 mm
A >> P	220 mm
F >> H	149 mm
Reset	Off

System - Tx/Rx

Frequency 1H	63.670263 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	5500.0 ms
Concatenations	2

Physio - Cardiac

Magn. preparation	Slice-sel. IR
TI	1930 ms
Fat suppr.	Fat sat.
Dark blood	Off
FoV read	220 mm
FoV phase	87.5 %
Phase resolution	100 %

Physio - Cardiac

Trajectory	Cartesian
------------	-----------

Physio - PACE

Resp. control	Off
Concatenations	2

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	7.8 ms
Bandwidth	287 Hz/Px

Sequence - Part 2

Define	Turbo factor
Echo trains per slice	9
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Normal
Gradient mode	Fast
WARP	Off
Red. EC sensitivity	Off
Turbo factor	25

Sequence - Assistant

Mode	Off
Allowed delay	30 s

\\USER\CRM AW\GOBrain\GOBrain_Aera_20ch\ep2d_diff_mddw_12_p2

TA: 1:21 PM: REF Voxel size: 1.5×1.5×5.0 mmPAT: 2 Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	31
Dist. factor	12 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	240 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	4500 ms
TE	77.0 ms
Averages	1
Concatenations	1
Filter	Dynamic Field Corr., Prescan Normalize
Coil elements	HE1-4

Contrast - Common

TR	4500 ms
TE	77.0 ms
MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	240 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	160
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
----------	--------

Resolution - iPAT

Accel. factor PE	2
Ref. lines PE	48
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On
Dynamic Field Corr.	On
Unfiltered images	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	31
Dist. factor	12 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	240 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	4500 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
AutoAlign	Head > Brain
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Special sat.	None

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm

System - Miscellaneous

MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	240 mm
R >> L	240 mm
F >> H	173 mm
Reset	Off

System - Tx/Rx

Frequency 1H	63.670263 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	4500 ms
Concatenations	1

Physio - PACE

Resp. control	Off
Concatenations	1

Diff - Neuro

Diffusion mode	MDDW
Diff. directions	12
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	800 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	On
ADC maps	On
FA maps	Off
Mosaic	On
Tensor	Off
Noise level	40

Diff - Body

Diffusion mode	MDDW
Diff. directions	12

Diff - Body

Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	800 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	On
ADC maps	On
Exponential ADC Maps	Off
FA maps	Off
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	40

Diff - Composing

Inline Composing	Off
Distortion Corr.	Off

Sequence - Part 1

Introduction	Off
Optimization	None
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.82 ms
Bandwidth	1420 Hz/Px

Sequence - Part 2

EPI factor	160
RF pulse type	Normal
Gradient mode	Fast

\\USER\CRM AW\GOBrain\GOBrain_Aera_20ch\ep2d_tra_hemo

TA: 6.1 s PM: REF Voxel size: 1.7×1.7×5.0 mmPAT: Off Rel. SNR: 1.00 : epfd

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	6120 ms
TE	75.0 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HE1-4

Contrast - Common

TR	6120 ms
TE	75.0 ms
MTC	Off
Flip angle	90 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	6120 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
AutoAlign	Head > Brain
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	220 mm
R >> L	220 mm
F >> H	149 mm
Reset	Off

System - Tx/Rx

Frequency 1H	63.670263 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	6120 ms
Concatenations	1

Perf

GBP	Off
PBP	Off
TTP	Off
relCBV	Off
relCBF	Off
relMTT	Off
relCBVCorr	Off
Measurements	1

Sequence - Part 1

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.77 ms
Bandwidth	1502 Hz/Px

Sequence - Part 2

EPI factor	128
RF pulse type	Normal
Gradient mode	Fast