

**Table of contents**

\USER

CRM AW

GOBrain

GOBrain\_Skyra\_20

AAHead\_Scout  
t1\_fld2d\_sag\_p2\_256  
t2\_tse\_tra\_p3\_256  
t2\_tse\_dark-fluid\_tra\_p2\_256  
ep2d\_diff\_mddw\_12\_p3  
ep2d\_tra\_hemo

\\USER\CRM AW\GOBrain\GOBrain\_Skyra\_20\AAHead\_Scout

TA: 0:14 PM: REF Voxel size: 1.6×1.6×1.6 mmPAT: 3 Rel. SNR: 1.00 : fl

**Properties**

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

**Routine**

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L0.0 P20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.15 ms
TE	1.37 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HE1-4

**Contrast - Common**

TR	3.15 ms
TE	1.37 ms
Flip angle	8 deg

**Contrast - Dynamic**

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

**Resolution - Common**

FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
Base resolution	160
Phase resolution	100 %
Slice resolution	69 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Trajectory	Cartesian

**Resolution - iPAT**

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	24
Accel. factor 3D	1

**Resolution - iPAT**

Reference scan mode	Integrated
---------------------	------------

**Resolution - Filter Image**

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

**Resolution - Filter Rawdata**

Raw filter	Off
Elliptical filter	Off

**Geometry - Common**

Slab group	1
Slabs	1
Dist. factor	20 %
Position	L0.0 P20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.15 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	1

**Geometry - AutoAlign**

Slab group	1
Position	L0.0 P20.0 H0.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

**System - Miscellaneous**

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off

**System - Miscellaneous**

Coil Select Mode	Off - AutoCoilSelect
------------------	----------------------

**System - Adjustments**

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

**System - Adjust Volume**

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

**System - Tx/Rx**

Frequency 1H	123.263740 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

**Physio - PACE**

Resp. control	Off
Concatenations	1

**Inline - Common**

Flip angle	8 deg
Measurements	1
Time to center	6.2 s

**Inline - Inline**

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

**Inline - MIP**

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

**Inline - Composing**

Inline Composing	Off
Distortion Corr.	Off

**Inline - MapIt**

Save original images	On
MapIt	None
Flip angle	8 deg
Measurements	1
Contrasts	1
TR	3.15 ms
TE	1.37 ms

**Sequence - Part 1**

Introduction	On
Dimension	3D
Asymmetric echo	Weak
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	540 Hz/Px

**Sequence - Part 2**

RF pulse type	Fast
Gradient mode	Normal
Excitation	Non-sel.
RF spoiling	On

**Sequence - Assistant**

Mode	Off
------	-----

\\USER\CRM AW\GOBrain\GOBrain\_Skyra\_20\t1\_fld2d\_sag\_p2\_256

TA: 0:41 PM: REF Voxel size: 0.9×0.9×4.0 mmPAT: 2 Rel. SNR: 1.00 : fl

**Properties**

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

**Resolution - iPAT**

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	48
Reference scan mode	Integrated

**Resolution - Filter Image**

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

**Routine**

Slice group	1
Slices	35
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Basis
Phase oversampling	40 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	240.0 ms
TE	2.46 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize, Elliptical filter
Coil elements	HE1-4

**Resolution - Filter Rawdata**

Raw filter	Off
Elliptical filter	On

**Geometry - Common**

Slice group	1
Slices	35
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	240.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

**Contrast - Common**

TR	240.0 ms
TE	2.46 ms
MTC	Off
Magn. preparation	None
Flip angle	80 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

**Geometry - AutoAlign**

Slice group	1
AutoAlign	Head > Basis
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

**Contrast - Dynamic**

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

**Geometry - Saturation**

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

**Resolution - Common**

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
Base resolution	256
Phase resolution	81 %
Phase partial Fourier	Off
Interpolation	Off

**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

**Geometry - Tim CT**

Tim CT mode	Off
Slices	35

**Geometry - Tim CT**

Slice thickness	4.0 mm
Dist. factor	20 %
FoV read	220 mm
FoV phase	100.0 %
Segments	1

**System - Miscellaneous**

Positioning mode	REF
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Basis
Coil Select Mode	Off - AutoCoilSelect

**System - Adjustments**

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

**System - Adjust Volume**

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

**System - Tx/Rx**

Frequency 1H	123.263740 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

**Physio - Signal1**

1st Signal/Mode	None
TR	240.0 ms
Concatenations	1
Segments	1

**Physio - Cardiac**

Tagging	None
Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	220 mm
FoV phase	100.0 %
Phase resolution	81 %

**Physio - PACE**

Resp. control	Off
---------------	-----

**Physio - PACE**

Concatenations	1
----------------	---

**Inline - Common**

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

**Inline - MIP**

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

**Inline - Soft Tissue**

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

**Inline - Composing**

Inline Composing	Off
Distortion Corr.	Off

**Inline - MapIt**

Save original images	On
MapIt	None
Flip angle	80 deg
Measurements	1
Contrasts	1
TR	240.0 ms
TE	2.46 ms

**Sequence - Part 1**

Introduction	Off
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Allowed
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Bandwidth	360 Hz/Px

**Sequence - Part 2**

Segments	1
Acoustic noise reduction	None
RF pulse type	Fast
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

**Sequence - Assistant**

Mode	Off
Allowed delay	40 s

\\USER\CRM AW\GOBrain\GOBrain\_Skyra\_20\t2\_tse\_tra\_p3\_256

TA: 1:02 PM: REF Voxel size: 0.9×0.9×5.0 mmPAT: 3 Rel. SNR: 1.00 : tse

**Properties**

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

**Routine**

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
TR	6200.0 ms
TE	78 ms
Averages	1
Concatenations	2
Filter	Distortion Corr.(2D), Prescan Normalize, Elliptical filter
Coil elements	HE1-4

**Contrast - Common**

TR	6200.0 ms
TE	78 ms
TD	0.0 ms
MTC	Off
Magn. preparation	None
Flip angle	120 deg
Fat suppr.	None
Water suppr.	None
Restore magn.	Off

**Contrast - Dynamic**

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

**Resolution - Common**

FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
Base resolution	256
Phase resolution	98 %
Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	Off

**Resolution - iPAT**

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	46
Reference scan mode	Integrated

**Resolution - Filter Image**

Image Filter	Off
Distortion Corr.	On
Mode	2D
Unfiltered images	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

**Resolution - Filter Rawdata**

Raw filter	Off
Elliptical filter	On

**Geometry - Common**

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
TR	6200.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	2

**Geometry - AutoAlign**

Slice group	1
AutoAlign	Head > Brain
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

**Geometry - Saturation**

Fat suppr.	None
Water suppr.	None
Restore magn.	Off
Special sat.	None

**Geometry - Navigator****Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

**Geometry - Tim CT**

Tim CT mode	Off
Slices	25
Slice thickness	5.0 mm
Dist. factor	20 %
FoV read	220 mm
FoV phase	87.5 %

**System - Miscellaneous**

Positioning mode	REF
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - AutoCoilSelect

**System - Adjustments**

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

**System - Adjust Volume**

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

**System - Tx/Rx**

Frequency 1H	123.263740 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

**Physio - Signal1**

1st Signal/Mode	None
TR	6200.0 ms
Concatenations	2

**Physio - Cardiac**

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	220 mm
FoV phase	87.5 %
Phase resolution	98 %
Trajectory	Cartesian

**Physio - PACE**

Resp. control	Off
---------------	-----

**Physio - PACE**

Concatenations	2
----------------	---

**Inline - Common**

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

**Inline - MIP**

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

**Inline - Composing**

Inline Composing	Off
Distortion Corr.	On
Mode	2D
Unfiltered images	Off

**Sequence - Part 1**

Introduction	Off
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	11.1 ms
Bandwidth	260 Hz/Px

**Sequence - Part 2**

Define	Turbo factor
Echo trains per slice	4
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Low SAR
Gradient mode	Fast
Hyperecho	On
WARP	Off
Red. EC sensitivity	Off
Turbo factor	26

**Sequence - Assistant**

Mode	Off
Allowed delay	60 s

\\USER\CRM AW\GOBrain\GOBrain\_Skyra\_20\t2\_tse\_dark-fluid\_tra\_p2\_256

TA: 1:52 PM: REF Voxel size: 0.9×0.9×5.0 mmPAT: 2 Rel. SNR: 1.00 : tir

**Properties**

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

**Routine**

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Phase oversampling	25 %
FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
TR	8000.0 ms
TE	119.0 ms
Averages	1
Concatenations	2
Filter	Prescan Normalize, Elliptical filter
Coil elements	HE1-4

**Contrast - Common**

TR	8000.0 ms
TE	119.0 ms
TD	0.0 ms
MTC	Off
Magn. preparation	Slice-sel. IR
TI	2370 ms
Flip angle	150 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong
Water suppr.	None
Restore magn.	Off
Freeze suppressed tissue	On

**Contrast - Dynamic**

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

**Resolution - Common**

FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
Base resolution	256
Phase resolution	98 %

**Resolution - Common**

Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	Off

**Resolution - iPAT**

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	26
Reference scan mode	Integrated

**Resolution - Filter Image**

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

**Resolution - Filter Rawdata**

Raw filter	Off
Elliptical filter	On

**Geometry - Common**

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	220 mm
FoV phase	87.5 %
Slice thickness	5.0 mm
TR	8000.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	2

**Geometry - AutoAlign**

Slice group	1
AutoAlign	Head > Brain
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	R >> L
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

**Geometry - Saturation**

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Water suppr.	None
Restore magn.	Off
Special sat.	None

**Geometry - Navigator**



**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

**Geometry - Tim CT**

Tim CT mode	Off
Slices	25
Slice thickness	5.0 mm
Dist. factor	20 %
FoV read	220 mm
FoV phase	87.5 %

**System - Miscellaneous**

Positioning mode	REF
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - AutoCoilSelect

**System - Adjustments**

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

**System - Adjust Volume**

Position	Isocenter
Orientation	Transversal
Rotation	90.00 deg
R >> L	193 mm
A >> P	220 mm
F >> H	149 mm
Reset	Off

**System - Tx/Rx**

Frequency 1H	123.263740 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

**Physio - Signal1**

1st Signal/Mode	None
TR	8000.0 ms
Concatenations	2

**Physio - Cardiac**

Magn. preparation	Slice-sel. IR
TI	2370 ms
Fat suppr.	Fat sat.
Dark blood	Off

**Physio - Cardiac**

FoV read	220 mm
FoV phase	87.5 %
Phase resolution	98 %
Trajectory	Cartesian

**Physio - PACE**

Resp. control	Off
Concatenations	2

**Inline - Common**

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

**Inline - MIP**

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

**Inline - Composing**

Inline Composing	Off
Distortion Corr.	Off

**Sequence - Part 1**

Introduction	Off
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	8.52 ms
Bandwidth	287 Hz/Px

**Sequence - Part 2**

Define	Turbo factor
Echo trains per slice	6
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Normal
Gradient mode	Fast
Hyperecho	Off
WARP	Off
Red. EC sensitivity	Off
Turbo factor	25

**Sequence - Assistant**

Mode	Off
Allowed delay	60 s

\\USER\CRM AW\GOBrain\GOBrain\_Skyra\_20\ep2d\_diff\_mddw\_12\_p3

TA: 1:16 PM: REF Voxel size: 1.5×1.5×5.0 mmPAT: 2 Rel. SNR: 1.00 : epse

**Properties**

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

**Routine**

Slice group	1
Slices	31
Dist. factor	12 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	240 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	4200 ms
TE	72.0 ms
Averages	1
Concatenations	1
Filter	Dynamic Field Corr., Prescan Normalize
Coil elements	HE1-4

**Contrast - Common**

TR	4200 ms
TE	72.0 ms
MTC	Off
Magn. preparation	None
Fat suppr.	Fat sat.
Fat sat. mode	Strong

**Contrast - Dynamic**

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

**Resolution - Common**

FoV read	240 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	160
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off

**Resolution - iPAT**

PAT mode	GRAPPA
----------	--------

**Resolution - iPAT**

Accel. factor PE	2
Ref. lines PE	38
Reference scan mode	EPI/separate

**Resolution - Filter Image**

Distortion Corr.	Off
Prescan Normalize	On
Dynamic Field Corr.	On
Unfiltered images	Off

**Resolution - Filter Rawdata**

Raw filter	Off
Elliptical filter	Off

**Geometry - Common**

Slice group	1
Slices	31
Dist. factor	12 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	240 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	4200 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

**Geometry - AutoAlign**

Slice group	1
AutoAlign	Head > Brain
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

**Geometry - Saturation**

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Special sat.	None

**Geometry - Navigator****Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

**System - Miscellaneous**

Positioning mode	REF
Table position	H
Table position	0 mm

**System - Miscellaneous**

MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - AutoCoilSelect

**System - Adjustments**

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

**System - Adjust Volume**

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	240 mm
R >> L	240 mm
F >> H	173 mm
Reset	Off

**System - Tx/Rx**

Frequency 1H	123.263740 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

**Physio - Signal1**

1st Signal/Mode	None
TR	4200 ms
Concatenations	1

**Physio - PACE**

Resp. control	Off
Concatenations	1

**Diff - Neuro**

Diffusion mode	MDDW
Diff. directions	12
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm <sup>2</sup>
b-value 2	800 s/mm <sup>2</sup>
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	On
ADC maps	On
FA maps	On
Mosaic	On
Tensor	On
Noise level	40

**Diff - Body**

Diffusion mode	MDDW
----------------	------

**Diff - Body**

Diff. directions	12
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm <sup>2</sup>
b-value 2	800 s/mm <sup>2</sup>
b-value 1	1
b-value 2	1
Diff. weighted images	On
Trace weighted images	On
ADC maps	On
Exponential ADC Maps	Off
FA maps	On
Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm <sup>2</sup>
Noise level	40

**Diff - Composing**

Inline Composing	Off
Distortion Corr.	Off

**Sequence - Part 1**

Introduction	Off
Optimization	None
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.94 ms
Bandwidth	1202 Hz/Px

**Sequence - Part 2**

EPI factor	160
RF pulse type	Normal
Gradient mode	Fast

\\USER\CRM AW\GOBrain\GOBrain\_Skyra\_20\ep2d\_tra\_hemo

TA: 6.1 s PM: REF Voxel size: 0.9×0.9×5.0 mmPAT: Off Rel. SNR: 1.00 : epfd

**Properties**

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

**Routine**

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	6120 ms
TE	30.0 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HE1-4

**Contrast - Common**

TR	6120 ms
TE	30.0 ms
MTC	Off
Flip angle	90 deg
Fat suppr.	Fat sat.

**Contrast - Dynamic**

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

**Resolution - Common**

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	On

**Resolution - iPAT**

PAT mode	None
----------	------

**Resolution - Filter Image**

Distortion Corr.	Off
Prescan Normalize	On

**Resolution - Filter Rawdata**

Raw filter	Off
Elliptical filter	Off
Hamming	Off

**Geometry - Common**

Slice group	1
Slices	25
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	6120 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

**Geometry - AutoAlign**

Slice group	1
AutoAlign	Head > Brain
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

**Geometry - Saturation**

Fat suppr.	Fat sat.
Special sat.	None

**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

**System - Miscellaneous**

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - AutoCoilSelect

**System - Adjustments**

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

**System - Adjust Volume**

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	220 mm
R >> L	220 mm
F >> H	149 mm
Reset	Off

**System - Tx/Rx**

Frequency 1H	123.263740 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

**Physio - Signal1**

1st Signal/Mode	None
TR	6120 ms
Concatenations	1

**Perf**

GBP	Off
PBP	Off
TTP	Off
relCBV	Off
relCBF	Off
relIMTT	Off
relCBVCorr	Off
Measurements	1

**Sequence - Part 1**

Introduction	Off
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.77 ms
Bandwidth	1502 Hz/Px

**Sequence - Part 2**

EPI factor	128
RF pulse type	Normal
Gradient mode	Fast