

Table of contents

\\USER				
	Knee			
		3D		
			CAIPI_SPACE_Knee	
				pd_space_sag_p4_iso t2_space_fs_sag_p4_iso

\\USER\Knee\3D\CAIPI_SPACE_Knee\pd_space_sag_p4_iso

TA: 4:08 PM: ISO Voxel size: 0.6×0.6×0.6 mmPAT: 4 Rel. SNR: 1.00 : spcR

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
Slice oversampling	23.1 %
Slices per slab	208
FoV read	160 mm
FoV phase	100.0 %
Slice thickness	0.63 mm
TR	900 ms
TE	28.0 ms
Averages	1.0
Concatenations	1
Filter	Distortion Corr.(2D), Prescan Normalize, Image Filter
Coil elements	15K

Contrast - Common

TR	900 ms
TE	28.0 ms
MTC	Off
Magn. preparation	None
Flip angle	120 deg
Fat suppr.	None
Blood suppr.	Off
Restore magn.	On

Contrast - Dynamic

Averages	1.0
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	160 mm
FoV phase	100.0 %
Slice thickness	0.63 mm
Base resolution	256
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	CAIPIRINHA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	2
Ref. lines 3D	24
Reordering Shift 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	On
! Intensity	Sharp
Edge Enhancement	1
Smoothing	1
Unfiltered images	Off
Distortion Corr.	On
Mode	2D
Unfiltered images	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	23.1 %
Slices per slab	208
FoV read	160 mm
FoV phase	100.0 %
Slice thickness	0.63 mm
TR	900 ms
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
Phase	0.0 mm
Read	0.0 mm
Shift	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Restore magn.	On
Special sat.	None

Geometry - Navigator

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	ISO
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	H >> F
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	0.00 deg
A >> P	160 mm
F >> H	160 mm
R >> L	132 mm
Reset	Off

System - Tx/Rx

Frequency 1H	63.679742 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
Trigger delay	0 ms
TR	900 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	160 mm
FoV phase	100.0 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	On
Mode	2D
Unfiltered images	Off

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	On
Reordering	Radial
Flow comp.	No
Echo spacing	5.58 ms
Bandwidth	425 Hz/Px

Sequence - Part 2

Echo train duration	301 ms
RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
Flip angle mode	Constant
Turbo factor	49

Sequence - Assistant

Allowed delay	0 s
---------------	-----

\\USER\Knee\3D\CAIPI_SPACE_Kneet2_space_fs_sag_p4_iso

TA: 5:59 PM: ISO Voxel size: 0.8×0.8×0.8 mmPAT: 4 Rel. SNR: 1.00 : spc

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	57 %
Slice oversampling	50.0 %
Slices per slab	160
FoV read	160 mm
FoV phase	100.0 %
Slice thickness	0.83 mm
TR	1050 ms
TE	109 ms
Averages	1.0
Concatenations	1
Filter	Distortion Corr.(2D), Prescan Normalize, Image Filter
Coil elements	15K

Contrast - Common

TR	1050 ms
TE	109 ms
MTC	Off
Magn. preparation	None
Fat suppr.	SPAIR
Fat sat. mode	Strong
Blood suppr.	Off
Restore magn.	Off

Contrast - Dynamic

Averages	1.0
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	160 mm
FoV phase	100.0 %
Slice thickness	0.83 mm
Base resolution	192
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Allowed
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	CAIPIRINHA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	2
Ref. lines 3D	24
Reordering Shift 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	On
! Intensity	Sharp
Edge Enhancement	1
Smoothing	1
Unfiltered images	Off
Distortion Corr.	On
Mode	2D
Unfiltered images	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	50.0 %
Slices per slab	160
FoV read	160 mm
FoV phase	100.0 %
Slice thickness	0.83 mm
TR	1050 ms
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
Phase	0.0 mm
Read	0.0 mm
Shift	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	SPAIR
Fat sat. mode	Strong
Restore magn.	Off
Special sat.	None

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table position	H
Table position	0 mm
Inline Composing	Off

System - Miscellaneous

Positioning mode	ISO
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	H >> F
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Sagittal
Rotation	0.00 deg
A >> P	160 mm
F >> H	160 mm
R >> L	133 mm
Reset	Off

System - Tx/Rx

Frequency 1H	63.679742 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
Trigger delay	0 ms
TR	1050 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	SPAIR
Dark blood	Off
FoV read	160 mm
FoV phase	100.0 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Inline Composing	Off
Distortion Corr.	On
Mode	2D
Unfiltered images	Off

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	On
Reordering	Linear
Flow comp.	No
Echo spacing	5.46 ms
Bandwidth	357 Hz/Px

Sequence - Part 2

Echo train duration	218 ms
RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
Flip angle mode	T2 var
Turbo factor	43

Sequence - Assistant

Allowed delay	0 s
---------------	-----