

IHE Integration Statement

syngo.via VB30A





IHE Integration Statement				Date	Apr 2018	
Vendor		Product Name		Version		
Siemens Healthineers		syngo.via	ngo.via VB30A			
This product implements Integration Profiles, Actor		ctions required in the IHE Technical tions listed below:	Framework	to support	the IHE	
Integration Profiles Implemented		Actors Implemented		Options Implemented		
Scheduled Workflow (SWF)	Image	Image Display		No options defined		
Workflow (SWF)	Evide	Evidence Creator			None	
Access to Radiology Information (ARI)	Image	Image Display			Multiple Sources	
Portable Data for Imaging		Portable Media Creator		None		
(PDI)		Portable Media Importer		No options defined		
	Displa	Display			No options defined	
Evidence Documents (ED) Evidei	Evidence Creator		No options defined		
Invoke Image Display (IID) Image	Image Display		No options defined		
Consistent Time (CT)		ne Client		None		
Internet address for v	endor's I	HE information: www.siemens.	com/ihe			
Links to Standards Co	nforman	ce Statements for the Implement	ation			
HL7 www.siemens.com/hl7						
DICOM <u>www.siemens.com/dicom</u>						
	Link	s to general information on IHE				
North America:		Europe:	Asia/Pacific:		c:	
http://ihe.net/IHE_Worldwide/		http://ihe.net/IHE_Worldwide/	http://ihe.net/IHE_Worldwide/		<u>'orldwide/</u>	

On account of certain regional limitations of sales rights and service availability, we cannot guarantee that all products included in this brochure are available through the Siemens sales organization worldwide. Availability and packaging may vary by country and are subject to change without prior notice.

Some/All of the features and products described herein may not be available in the United States or other countries.

The information in this document contains general technical descriptions of specifications and options as well as standard and optional features that do not always have to be present in individual cases.

Siemens reserves the right to modify the design, packaging, specifications and options described herein without prior notice. Please contact your local Siemens sales representative for the most current information.

In the interest of complying with legal requirements concerning the environmental compatibility of our products (protection of natural resources and waste conservation), we recycle certain components. Using the same extensive quality assurance measures as for factorynew components, we guarantee the quality of these recycled components.

Note: Any technical data contained in this document may vary within defined tolerances. Original images always lose a certain amount of detail when reproduced.

Caution: Federal law restricts this device to sale by or on the order of a physician.

Siemens Healthineers Headquarters

Siemens Healthcare GmbH Henkestr. 127 91052 Erlangen Germany Phone +49 9131 84-0 siemens.com/healthineers