



ACUSON Sequoia Ultrasound System

Environmental Product Declaration

siemens-healthineers.com/sequoia





Taking ultrasound to new heights

Going back to the future with the new ACUSON Sequoia. Variability has hampered ultrasound's potential to expand precision medicine. Today health systems struggle with rising costs and varying quality. Ultrasound is one of the most widely-used and readily available imaging modality.¹ From screening and diagnosis, to therapy planning and monitoring, ultrasound has the potential to expand precision medicine throughout the patient care continuum. To achieve this goal, ultrasound needs to address biological, technological and user variabilities.

Key product features

- OLED monitor with high dynamic range display for superior detail resolution
- 15.6" touch display featuring context sensitive lighting for easy work-flow solutions
- Floating control panel design with 180 degree left/right swivel
- Small footprint, lightweight design for easy portable use


¹Regional Diagnostic Radiology: <http://www.rdradiology.com/radiology-services-modalities.html>

ACUSON Sequoia Ultrasound System

BioAcoustic imaging Technology

It all starts with the way we generate, track, transmit and receive the ultrasound signals. With the new ACUSON Sequoia powered by BioAcoustic technology, each individual component is assembled to accurately track the ultrasound signal throughout the signal path. From the power supplies to our receivers and graphics processing unit; to the compact-pinless connectors, to the transducer lens, we aimed to preserve the signal acoustic fidelity.

BioAcoustic imaging technology adapts the ultrasound signal in real time, to compensate for loss of energy to each patients biological characteristics. The new ACUSON Sequoia has 10x higher acoustic fidelity than conventional ultrasound systems.¹



Environmental benefits

- Uses 33% less electricity during scanning²
- Weighs 20% less²
- Fast boot-up reduces electricity usage

Customer benefits

- InFocus Coherent Image Formation – fully focused imaging from near field to far field in all applications and modes.
- DAX (Deep Abdominal Transducer) designed to go deep, with up to 40 cm of diagnostic imaging, whilst reducing the necessary force required to scan large patients.
- Gesture detecting transducer technology allows the user to activate the transducer by touch rather than repetitive key strokes and reaches.
- AI-enabled and automation technologies to reduce user key strokes and improve exam quality.

¹Compared to ACUSON S3000 ultrasound system

²Compared to ACUSON SC2000 ultrasound system

Environmental Management System

Siemens Healthineers gives high priority to achieving excellence in Environmental Protection, Health Management and Safety (EHS).

Across the globe, Siemens Healthineers has implemented a consistent EHS management system. It lays the foundation for the continuous improvement of our performance in these areas, and regular auditing assures our conformance.

As a result of this consistent approach, Siemens Healthineers is considered one organization and is certified in accordance with ISO 14001 and OHSAS 18001.

Environmental Product Design



Material supply:
From natural resources to delivery of semi-finished products



Production/delivery:
From production of components to operation startup by the customer



Use/maintenance:
Includes daily use by our customers as well as maintenance



End-of-life:
From disassembly at the customer site, through material and energy recycling


Siemens Healthineers considers environmental aspects in all phases of the product life cycle, including material supply, production/delivery, use/maintenance and end of life.

Our product design procedure fulfills the requirements of IEC 60601-1-9:2007 + A1 2013 Medical electrical equipment Part 1-9: General requirements for basic safety and essential performance – Collateral Standard: Requirements for environmentally conscious design.

This standard supports the effort to improve the environmental performance of our products.

Cumulative Energy Demand

Energy consumption is the most important environmental characteristic of medical devices. This is why we use the Cumulative Energy Demand to assess environmental performance. Cumulative Energy Demand is the total primary energy¹ that is necessary to produce, use and dispose of a device – including all transportation. Our medical devices can be recycled almost completely for materials or energy. With an appropriate end-of-life treatment it is possible to return up to -3 MWh in the form of secondary raw materials or thermal energy to the economic cycle.




¹Primary energy is the energy contained in natural resources prior to undergoing any man made conversions (e.g. oil, solar).

²Based on 10 hours of use (5 hours scanning, 5 hours freeze) 5 days a week, 10 years usage.

Product Materials

ACUSON Sequoia Ultrasound System is mainly built out of metals. This ensures a high degree of recyclability.

Total weight: approx. 125 kg




Packaging Materials

It is our goal to minimize our packaging material and reduce the packaging waste by reusing and recycling it.

Nearly all the packaging is recyclable either for reuse of the materials or for energy recovery. Only an insignificant amount (~ 1%) has to be sent to landfill.

Total weight: 43.5 kg



Product Take Back

Most of the materials used to produce the ACUSON Sequoia system are recyclable. Over 96% (by weight) can be recycled for material content and 4% for energy. Disassembly instructions for disposal and recycling are available for our products.



Operating Data

Heat emissions of the device

• Off	12.7 W (100 V / 60 Hz), 12.6 W (115 V / 60 Hz), 13.4 W (230 V / 50 Hz) (approx.)
• Freeze	358 W (100 V / 60 Hz), 357 W (115 V / 60 Hz), 355 W (230 V / 50 Hz) (approx.)
• Scanning	386 W (100 V / 60 Hz), 386 W (115 V / 60 Hz), 383 W (230 V / 50 Hz) (approx.)
• Hibernate/standby	12.3 W (100 V / 60 Hz), 12.3 W (115 V / 60 Hz), 13.4 W (230 V / 50 Hz) (approx.)

Allowed ambient temperature

• During operation	+10 °C to +40 °C (system without a printer) +10 °C to +35 °C (system with a printer)
• During storage or transportation	-10 °C to +60 °C (system without a printer) -10 °C to +60 °C (system with a printer)

Allowed relative humidity

• During operation	20–80% non-condensing
• During storage or transportation	10–95% non-condensing

Typical power consumption

• Off	12.7 W (100 V / 60 Hz), 12.6 W (115 V / 60 Hz), 13.4 W (230 V / 50 Hz)
• Freeze	358 W (100 V / 60 Hz), 357 W (115 V / 60 Hz), 355 W (230 V / 50 Hz)
• Scanning	386 W (100 V / 60 Hz), 386 W (115 V / 60 Hz), 383 W (230 V / 50 Hz)
• Hibernate/standby	12.3 W (100 V / 60 Hz), 12.3 W (115 V / 60 Hz), 13.4 W (230 V / 50 Hz)

Power-on time	2 minutes 30 seconds
---------------	----------------------

Power-off time	19.8 seconds
----------------	--------------

Technical Specifications

Interface for heat recovery	None
-----------------------------	------

Possible type of cooling	Air
--------------------------	-----

Complete switch-off is possible	Yes
---------------------------------	-----

Device is adjustable for the user in terms of height	Yes
--	-----

Uniform operating symbols for device families	Yes
---	-----

Radiation

This product produces no ionizing radiation

Electromagnetic Fields

Measures/techniques to minimize the exposure to electromagnetic fields	<ul style="list-style-type: none"> • complies to EN 55011/CSPR11 • power filtering • electromagnetic shielding • cable shielding • grounded metallic components
--	--

Replacement Parts and Consumables

- Needle guides
- Brackets for needle guides

Disposal/Substance Information

End of Life concept	Yes
Recycling information	Yes
List of hazardous substance (not contained in device)	Yes

Cleaning

Incompatible cleaning processes:

Total device

- Do not clean the system with chlorinated or aromatic solvents, acidic or basic solutions, isopropyl alcohol or strong cleaners such as ammoniated products.
- Do not use spray cleaners on the ultrasound system.
- Do not pour any fluid onto the ultrasound system surfaces.
- Iodine and betadine may discolor the control panel.
- Use only approved disinfectant wipes on the surfaces of the ultrasound system and accessories.

Restrictions for particular device components

- See Instructions for Use for cleaning and disinfecting transducers.
- Use only approved disinfectants on the holders for transducers and coupling gel holder.
- Never use gauze pads, lint-free cloths, or solutions to clean the transducer ports.
- Do not use disinfectant wipes on the glass surfaces of the touch screen or monitor. Carefully follow the instruction to clean the glass surfaces of the touch screen and monitor.
- Do not immerse the gel warmer in water or any solution.
- Use an approved disinfectant wipe to disinfect the ring, trackball, and trackball assembly.

Suitability of device for sterile areas No

Size of the surface to be cleaned 0.22 m²

Please refer to the dedicated user and reference manuals for system and components for a detailed list of approved and not approved cleaning substances and further instructions.

Further Ecologically Relevant Information

Elements of instructions are:

- Recommendations for saving energy Yes
- Recommendations for efficient cleaning Yes
- Recommendations for appropriate use of consumables Yes



On account of certain regional limitations of sales rights and service availability, we cannot guarantee that all products included in this brochure are available through the Siemens Healthineers sales organization worldwide. Availability and packaging may vary by country and is subject to change without prior notice. Some/All of the features and products described herein may not be available in the United States.

The information in this document contains general technical descriptions of specifications and options as well as standard and optional features which do not always have to be present in individual cases.

Siemens Healthineers reserves the right to modify the design, packaging, specifications and options described herein without prior notice. Please contact your local Siemens Healthineers sales representative for the most current information.

Note: Any technical data contained in this document may vary within defined tolerances. Original images always lose a certain amount of detail when reproduced.

The statements by Siemens Healthineers' customers described herein are based on results that were achieved in the customer's unique setting. Since there is no "typical" hospital and many variables exist (e.g., hospital size, case mix, level of IT adoption) there can be no guarantee that other customers will achieve the same results.

ACUSON Sequoia and BioAcoustic, are trademarks of Siemens Medical Solutions USA, Inc.

Siemens Healthineers Headquarters

Siemens Healthcare GmbH
Henkestr. 127
91052 Erlangen, Germany
Phone: +49 9131 84-0
siemens-healthineers.com

Legal Manufacturer

Siemens Medical Solutions USA, Inc.
Ultrasound
22010 S.E. 51st Street
Issaquah, WA 98029, USA
Phone: 1-888-826-9702
siemens-healthineers.com/ultrasound