



IHE Integration Statement

syngo.via

VB50A

IHE Integration Statement		Date	June 2020
Vendor	Product Name		Version
Siemens Healthineers	syngo.via		VB50A
This product implements all transactions required in the IHE Technical Framework to support the IHE Integration Profiles, Actors and Options listed below:			
Integration Profiles Implemented	Actors Implemented		Options Implemented
Scheduled Workflow (SWF)	Image Display	No options defined	
	Evidence Creator	None	
Access to Radiology Information (ARI)	Image Display	Multiple Sources	
Portable Data for Imaging (PDI)	Portable Media Creator	None	
	Portable Media Importer	None	
	Display	No options defined	
Evidence Documents (ED)	Evidence Creator	No options defined	
Invoke Image Display (IID)	Image Display	No options defined	
Consistent Time (CT)	Time Client	None	
Internet address for vendor’s IHE information: www.siemens.com/ihe			
Links to Standards Conformance Statements for the Implementation			
HL7	www.siemens.com/hl7		
DICOM	www.siemens.com/dicom		
Links to general information on IHE			
North America: http://ihe.net/IHE_Worldwide/	Europe: http://ihe.net/IHE_Worldwide/	Asia/Pacific: http://ihe.net/IHE_Worldwide/	

On account of certain regional limitations of sales rights and service availability, we cannot guarantee that all products included in this brochure are available through the Siemens sales organization world-wide. Availability and packaging may vary by country and are subject to change without prior notice.

Some/All of the features and products described herein may not be available in the United States or other countries.

The information in this document contains general technical descriptions of specifications and options as well as standard and optional features that do not always have to be present in individual cases.

Siemens reserves the right to modify the design, packaging, specifications and options described herein without prior notice. Please contact your local Siemens sales representative for the most current information.

In the interest of complying with legal requirements concerning the environmental compatibility of our products (protection of natural resources and waste conservation), we recycle certain components. Using the same extensive quality assurance measures as for factory-new components, we guarantee the quality of these recycled components.

Note: Any technical data contained in this document may vary within defined tolerances. Original images always lose a certain amount of detail when reproduced.

Caution: Federal law restricts this device to sale by or on the order of a physician.

Siemens Healthineers Headquarters
Siemens Healthcare GmbH
Henkestr. 127
91052 Erlangen
Germany
Phone +49 9131 84-0
[siemens.com/healthineers](https://www.siemens.com/healthineers)