



epoc Blood Analysis System with epoc NXS Host

Summary of Analytical Methods and Performance

siemens-healthineers.com/epocnxs



epoc® Blood Analysis System with epoc® NXS Host **Summary of Analytical Methods** **and Performance**





Table of Contents

Performance Data	4
Glossary	4
Methodologies	4
pH	6
$p\text{CO}_2$	8
$p\text{O}_2$	10
TCO_2	12
Sodium	14
Potassium	16
Ionized Calcium	18
Chloride	20
Hematocrit	22
Glucose	24
Lactate	26
BUN	28
Creatinine	30

Performance data

The data summarized here are compiled from user performance verifications of the epoc® Blood Analysis System, performed as part of the implementation process.

Precision

The precision data provided for each analyte are the pooled averages of the precision data from performance verifications from 1–12 user sites.

Method comparison

Method comparison studies were performed by individuals who were thoroughly familiar with the operation, maintenance, and control of both the epoc system and comparative method systems before starting. Testing was performed at all sites using blood collected in either blood gas syringes or in green-top evacuated tubes. Some samples were spiked with concentrated solutions to create samples with concentrations throughout the reportable range of each analyte. Each plot included in this summary is from 1–3 sites and is representative of the comparison of the epoc Blood Analysis System to each instrument.

Glossary

Accuracy is how close a result is to its true value.

Precision is reproducibility—how closely multiple results obtained from the same sample agree with each other.

n is the number of data points included in the data set.

x represents the comparison method in regression analysis.

y represents the test method in regression analysis.

Slope describes the angle of the line that provides the best fit of the test and comparison results. A perfect slope would be 1.00. Deviations from 1.00 are an indication of proportional systematic error.¹

Intercept (int't) or y-intercept describes where the line of best fit intersects the y-axis. The y-intercept should be an indication of constant systematic error.¹

Sy.x describes the scatter of the data around the line of best fit. It provides an estimate of the random error between the methods and includes both the imprecision of the test and comparison methods, as well as possible matrix effects that vary from one sample to another. Sy.x will never be 0 because both methods have some imprecision.¹

r or **correlation coefficient** describes how closely the results between the two methods change together. The lower the r value, the more scatter there is in the data. The main use of r is to help assess the reliability of the regression data—r should never be used as an indicator of method acceptability.¹

Methodologies

pH is measured by potentiometry using a pH-selective membrane electrode. The concentration of hydrogen ions is obtained from the measured potential using the Nernst equation.

pCO₂ is measured by potentiometry using a membrane-covered pH-sensing electrode. The electrode voltage is proportional to the dissolved carbon dioxide concentration through the Nernst equation.

pO₂ is measured by amperometry using a membrane-covered oxygen-sensing cathode electrode. The oxygen reduction current is proportional to the dissolved oxygen concentration.

TCO₂ is measured based on a modified Henderson-Hasselbalch equation, using pH and pCO₂,² and calibrated to match the International Federation of Clinical Chemistry (IFCC) Reference Measurement Procedure for Total Carbon Dioxide.³ Therefore, it is metrologically traceable to the IFCC TCO₂ reference method.³

Sodium is measured by potentiometry using an ion-selective membrane electrode. The concentration of sodium ions is obtained from the measured potential using the Nernst equation. The epoc sodium measurement is an undiluted (direct) method. Values may differ from those obtained by dilutional (indirect) methods.


Potassium is measured by potentiometry using an ion-selective membrane electrode. The concentration of potassium ions is obtained from the measured potential using the Nernst equation. The epoc potassium measurement is an undiluted (direct) method. Values may differ from those obtained by dilutional (indirect) methods.

Ionized calcium is measured by potentiometry using an ion-selective membrane electrode. The concentration of calcium ions is obtained from the measured potential using the Nernst equation.


Chloride is measured by potentiometry using an ion-selective membrane electrode. The concentration of chloride ions is obtained from the measured potential using the Nernst equation.

Hematocrit is measured by AC conductometry using two gold electrodes. The conductance of the blood sample in the fluidic path between the two electrodes, after correction for variable plasma conductivity through the measurement of sodium and potassium concentration, is inversely proportional to the hematocrit value.

Glucose is measured by amperometry. The sensor comprises an immobilized enzyme first layer coated onto a gold electrode of the electrode module, with a diffusion barrier second layer. The glucose oxidase enzyme is employed to convert glucose to hydrogen peroxide:




and then uses an amperometric sensor to detect the enzymatically produced hydrogen peroxide. Peroxide detection is by redox-mediated (ABTS [2,2'-azino-bis 3-ethylbenzothiazoline-6-sulfonic acid] diammonium salt) horseradish peroxidase (HRP)-catalyzed reduction on a gold electrode.




The reduction current is proportional to the concentration of glucose in the test fluid. The epoc glucose result is reported as plasma-equivalent glucose concentration.

Lactate is measured by amperometry. The sensor comprises an immobilized enzyme first layer coated onto a gold electrode of the electrode module, with a diffusion barrier second layer. The lactate oxidase enzyme is employed to convert lactate to hydrogen peroxide:




and then uses an amperometric sensor to detect the enzymatically produced hydrogen peroxide. Peroxide detection is by redox-mediated (ABTS [2,2'-azino-bis 3-ethylbenzothiazoline-6-sulfonic acid] diammonium salt) horseradish peroxidase (HRP)-catalyzed reduction on a gold electrode.



The reduction current is proportional to the concentration of lactate in the test fluid.


BUN/Urea is measured by potentiometry using an ammonium ion-selective electrode coated onto a gold electrode, covered with an enzymatic membrane second layer. The urease enzyme is employed to convert urea to ammonium ions:




and then uses a potentiometric ion-selective electrode to detect the enzymatically produced ammonium ion. The concentration of ammonium ions is obtained from the measured potential using the Nernst equation.

Creatinine is measured by amperometry. Each creatinine sensor is a three-layer enzyme electrode comprising a first immobilized enzyme creatinine-conversion underlayer coated onto a gold electrode, a second immobilized enzyme creatine screening layer, and a third diffusion barrier layer.

The creatinine electrode underlayer contains the enzymes creatinine amidohydrolase, creatine amidinohydrolase, and sarcosine oxidase, which convert creatinine to hydrogen peroxide in an enzyme product cascade:



and then uses the underlying gold electrode to detect the enzymatically produced hydrogen peroxide. Peroxide detection is by redox-mediated horseradish peroxidase (HRP)-catalyzed reduction.



The reduction current is proportional to the concentration of creatinine in the test fluid.

References:


1. Westgard JO. Basic method evaluation. 3rd ed. Madison, WI (USA): Westgard QC, Inc.; 2008 p. 77–78.
2. Maas AH, Rispens P, Siggaard-Andersen O, Zijlstra WG. On the reliability of Henderson-Hasselbalch equation in routine clinical acid-base chemistry. Ann Clin Biochem. 1984;21:26-39.
3. International Federation of Clinical Chemistry and Laboratory Medicine. IFCC reference measurement procedure for substance concentration determination of total carbon dioxide in blood, plasma or serum. Clin Chem Lab Med. 2001;39(3).

pH Method Comparison

pH				
Precision	n	Mean	SD	%CV
Level 1	24	7.052	0.009	0.13%
Level 3	25	7.646	0.007	0.09%

pH


X: Abbott I-STAT System
Y: epoc System



n = 41
slope = 1.049
int't. = -0.356
Sy.x = 0.016
r = 0.997

pH


X: RAPIDLab® 1265 Blood Gas System by Siemens Healthineers
Y: epoc System



n = 25
slope = 0.990
int't. = 0.082
Sy.x = 0.013
r = 0.998

pH


X: Radiometer ABL 700 System
Y: epoc System



n = 42
slope = 1.015
int't. = -0.097
Sy.x = 0.015
r = 0.998

pH

X: GEM PREMIER 3000 System
Y: epoc System




n = 32
slope = 0.923
int't. = 0.566
Sy.x = 0.010
r = 0.992

pH

X: Nova Biomedical CRITICAL CARE XPRESS System

Y: epoc System



$n = 43$

slope = 0.996

int't. = -0.004


Sy.x = 0.017

$r = 0.982$

pH

X: IRMA TRUPOINT System

Y: epoc System



$n = 33$

slope = 1.117

int't. = -0.865

Sy.x = 0.010


$r = 0.993$

pCO₂ Method Comparison

pCO ₂ mmHg				
Precision	n	Mean	SD	%CV
Level 1	24	67.2	2.30	3.41%
Level 3	25	20.8	0.68	3.25%

pCO₂


X: Abbott I-STAT System
Y: epoc System



n = 41
slope = 1.058
int't. = -4.60
Sy.x = 2.03
r = 0.996

pCO₂


X: RAPIDLab 1265 Blood Gas System by Siemens Healthineers
Y: epoc System



n = 25
slope = 1.000
int't. = -0.91
Sy.x = 1.24
r = 0.999

pCO₂


X: Radiometer ABL 700 System
Y: epoc System



n = 26
slope = 0.977
int't. = -0.24
Sy.x = 1.63
r = 0.995

pCO₂

X: IL GEM PREMIER 3000 System
Y: epoc System




n = 52
slope = 1.002
int't. = -0.34
Sy.x = 2.47
r = 0.995

pCO₂

X: Nova Biomedical CRITICAL CARE XPRESS System

Y: epoc System



n = 46

slope = 1.006

int't. = 2.86


Sy.x = 2.88

r = 0.975

pCO₂

X: IRMA TRUPOINT System

Y: epoc System



n = 32

slope = 1.047

int't. = -2.49

Sy.x = 1.56


r = 0.979

pO₂ Method Comparison

pO ₂ mmHg				
Precision	n	Mean	SD	%CV
Level 1	24	63.7	4.46	7.00%
Level 3	25	185.6	6.46	3.48%

pO₂


X: Abbott I-STAT System
Y: epoc System



n = 42
slope = 0.949
int't. = 7.86
Sy.x = 4.78
r = 0.997

pO₂


X: RAPIDLab 1265 Blood Gas System by Siemens Healthineers
Y: epoc System



n = 24
slope = 1.018
int't. = 3.64
Sy.x = 4.04
r = 0.998

pO₂


X: Radiometer ABL 700 System
Y: epoc System



n = 51
slope = 0.919
int't. = 9.01
Sy.x = 5.80
r = 0.995

pO₂

X: IL GEM PREMIER 3000 System
Y: epoc System




n = 32
slope = 0.947
int't. = 14.20
Sy.x = 8.50
r = 0.987

pO_2

X: Nova Biomedical CRITICAL CARE XPRESS System

Y: epoc System



$n = 43$

slope = 0.900

int't. = 11.32


Sy.x = 7.30

$r = 0.997$

pO_2

X: IRMA TRUPOINT System

Y: epoc System



$n = 31$

slope = 1.047

int't. = -6.60

Sy.x = 5.13


$r = 0.971$

TCO₂ Method Comparison

TCO ₂ mmol/L				
Precision	n	Mean	SD	%CV
Level 1	136	18.7	0.23	1.2%
Level 3	132	30.8	0.54	1.7%

TCO₂


X: Abbott ARCHITECT System
Y: epoc System



n = 86
slope = 0.98
int't. = 2.2
Sy.x = 1.17
r = 0.973

TCO₂


X: Dimension® EXL™ Integrated Chemistry System by Siemens Healthineers
Y: epoc System



n = 65
slope = 1.05
int't. = -0.8
Sy.x = 1.17
r = 0.974

TCO₂


X: Beckman Coulter DXC System
Y: epoc System



n = 40
slope = 1.12
int't. = -2.1
Sy.x = 0.58
r = 0.989

TCO₂

X: Dimension Vista® Intelligent Lab System by Siemens Healthineers
Y: epoc System




n = 144
slope = 1.18
int't. = -3.7
Sy.x = 1.0
r = 0.977

TCO₂

X: Roche COBAS System

Y: epoc System




n = 80
slope = 1.02
int't. = 1.2
Sy.x = 1.04
r = 0.981

Sodium Method Comparison

Sodium mmol/L				
Precision	n	Mean	SD	%CV
Level 1	27	113	0.82	0.73%
Level 3	27	166	1.07	0.64%

Na⁺


X: Abbott I-STAT System
Y: epoc System



n = 63
slope = 0.927
int't. = 10.19
Sy.x = 1.82
r = 0.982

Na⁺


X: RAPIDLab 1265 Blood Gas System by Siemens Healthineers
Y: epoc System



n = 26
slope = 1.057
int't. = -5.30
Sy.x = 2.77
r = 0.922

Na⁺


X: Radiometer ABL 700 System
Y: epoc System



n = 26
slope = 1.010
int't. = -0.01
Sy.x = 2.55
r = 0.987

Na⁺


X: IL GEM PREMIER 3000 System
Y: epoc System



n = 58
slope = 1.000
int't. = 1.42
Sy.x = 1.05
r = 0.919

Na⁺


X: Nova Biomedical PHOX System
Y: epoc System



n = 43
slope = 0.944
int't. = 8.38
Sy.x = 2.18
r = 0.939

Na⁺


X: Ortho Clinical Laboratories VITROS System
Y: epoc System



n = 35
slope = 0.947
int't. = 6.70
Sy.x = 1.25
r = 0.871

Na⁺


X: Beckman Coulter DXC System
Y: epoc System



n = 25
slope = 0.975
int't. = 4.49
Sy.x = 2.00
r = 0.981

Na⁺

X: Dimension Integrated Chemistry System by Siemens Healthineers
Y: epoc System




n = 36
slope = 1.043
int't. = -4.15
Sy.x = 0.77
r = 0.994

Potassium Method Comparison

Potassium mmol/L				
Precision	n	Mean	SD	%CV
Level 1	27	2.1	0.043	2.04%
Level 3	27	6.3	0.075	1.20%

K⁺


X: Abbott I-STAT System
Y: epoc System



n = 38
slope = 0.980
int't. = 0.07
Sy.x = 0.099
r = 0.997

K⁺


X: RAPIDLab 1265 Blood Gas System by Siemens Healthineers
Y: epoc System



n = 26
slope = 1.019
int't. = -0.08
Sy.x = 0.141
r = 0.995

K⁺


X: Radiometer ABL 700 System
Y: epoc System



n = 26
slope = 1.023
int't. = -0.11
Sy.x = 0.082
r = 0.999

K⁺


X: IL GEM PREMIER 3000 System
Y: epoc System



n = 31
slope = 0.959
int't. = 0.13
Sy.x = 0.090
r = 0.995

K⁺


X: Nova Biomedical PHOX System
Y: epoc System



n = 43
slope = 1.042
int't. = -0.18
Sy.x = 0.122
r = 0.995

K⁺


X: Ortho Clinical Laboratories VITROS System
Y: epoc System



n = 54
slope = 0.965
int't. = -0.07
Sy.x = 0.072
r = 0.985

K⁺


X: Beckman Coulter DXC System
Y: epoc System



n = 26
slope = 0.991
int't. = 0.19
Sy.x = 0.063
r = 0.998

K⁺

X: Dimension Integrated Chemistry System by Siemens Healthineers
Y: epoc System




n = 43
slope = 0.948
int't. = 0.13
Sy.x = 0.101
r = 0.997

Ionized Calcium Method Comparison

Ionized Calcium mmol/L				
Precision	n	Mean	SD	%CV
Level 1	26	1.53	0.019	1.25%
Level 3	27	0.67	0.009	1.40%

Ca⁺⁺


X: Abbott I-STAT System
Y: epoc System



n = 39
slope = 0.997
int't. = 0.00
Sy.x = 0.025
r = 0.991

Ca⁺⁺


X: RAPIDLab 1265 Blood Gas System by Siemens Healthineers
Y: epoc System



n = 44
slope = 0.960
int't. = 0.04
Sy.x = 0.047
r = 0.969

Ca⁺⁺


X: Radiometer ABL 700 System
Y: epoc System



n = 25
slope = 1.004
int't. = -0.05
Sy.x = 0.035
r = 0.997

Ca⁺⁺

X: IL GEM PREMIER 3000 System
Y: epoc System




n = 31
slope = 0.979
int't. = 0.06
Sy.x = 0.027
r = 0.979

Ca⁺⁺

X: Nova Biomedical PHOX System

Y: epoc System



n = 43

slope = 0.986

int't. = 0.00

Sy.x = 0.039


r = 0.994

Chloride Method Comparison

Chloride mmol/L				
Precision	n	Mean	SD	%CV
Level 1	20	76	0.53	0.69%
Level 3	20	125	0.94	0.76%

Chloride


X: Abbott I-STAT System
Y: epoc System



n = 64
slope = 0.989
int't. = -0.525
Sy.x = 1.033
r = 0.995

Chloride


X: ADVIA® Clinical Chemistry System by Siemens Healthineers
Y: epoc System



n = 53
slope = 0.981
int't. = 1.084
Sy.x = 1.773
r = 0.985

Chloride


X: Roche COBAS System
Y: epoc System



n = 50
slope = 0.982
int't. = 5.032
Sy.x = 1.250
r = 0.990

Chloride

X: Beckman Coulter DXC System
Y: epoc System




n = 63
slope = 0.990
int't. = 1.611
Sy.x = 1.670
r = 0.982

Chloride

X: Radiometer ABL 800 System

Y: epoc System



$n = 56$

slope = 1.040

int't. = -4.866

Sy.x = 0.545


$r = 0.995$

Hematocrit Method Comparison

Hematocrit %PCV				
Precision	n	Mean	SD	%CV
Level 1	26	25	0.56	2.28%
Level 3	26	44	1.16	2.61%

Hct


X: Radiometer ABL 825 System
Y: epoc System



n = 38
slope = 0.996
int't. = -0.4
Sy.x = 1.81
r = 0.982

Hct


X: RAPIDLab 1265 Blood Gas System by Siemens Healthineers
Y: epoc System



n = 23
slope = 1.051
int't. = -4.0
Sy.x = 2.61
r = 0.971

Hct


X: Abbott I-STAT System
Y: epoc System



n = 29
slope = 0.944
int't. = 2.2
Sy.x = 1.40
r = 0.991

Hct


X: IL GEM PREMIER 3000 System
Y: epoc System



n = 57
slope = 1.037
int't. = -2.8
Sy.x = 2.83
r = 0.920

Hct


X: Nova Biomedical PHOX System
Y: epoc System



n = 34
slope = 1.052
int't. = -3.6
Sy.x = 1.76
r = 0.986

Hct


X: Sysmex XE System
Y: epoc System



n = 18
slope = 0.983
int't. = -0.4
Sy.x = 1.96
r = 0.971

Hct


X: Beckman Coulter LH System
Y: epoc System



n = 29
slope = 1.067
int't. = -0.3
Sy.x = 1.86
r = 0.984

Hct

X: Microcentrifugation (spun)
Y: epoc System




n = 63
slope = 0.963
int't. = 0.9
Sy.x = 2.01
r = 0.970

Glucose Method Comparison

Glucose mg/dL				
Precision	n	Mean	SD	%CV
Level 1	27	41.9	1.24	2.96%
Level 3	27	278	6.84	2.46%

Glucose


X: Abbott I-STAT System
Y: epoc System



n = 41
slope = 1.015
int't. = 1.8
Sy.x = 5.59
r = 0.999

Glucose


X: RAPIDLab 1265 Blood Gas System by Siemens Healthineers
Y: epoc System



n = 22
slope = 1.052
int't. = 4.0
Sy.x = 15.75
r = 0.990

Glucose


X: Radiometer ABL 700 System
Y: epoc System



n = 24
slope = 1.048
int't. = -1.7
Sy.x = 5.49
r = 0.995

Glucose


X: IL GEM PREMIER 3000 System
Y: epoc System



n = 31
slope = 1.042
int't. = 11.9
Sy.x = 11.07
r = 0.989

Glucose


X: Nova Biomedical CRITICAL CARE XPRESS System
Y: epoc System



n = 44
slope = 1.021
int't. = -4.7
Sy.x = 6.43
r = 0.994

Glucose


X: Ortho Clinical Laboratories VITROS System
Y: epoc System



n = 41
slope = 1.018
int't. = 0.8
Sy.x = 6.82
r = 0.998

Glucose


X: Beckman Coulter DXC System
Y: epoc System



n = 24
slope = 1.057
int't. = -10.5
Sy.x = 7.71
r = 0.996

Glucose

X: Dimension Integrated Chemistry System by Siemens Healthineers
Y: epoc System




n = 43
slope = 1.016
int't. = -2.7
Sy.x = 7.49
r = 0.997

Lactate Method Comparison

Lactate mmol/L				
Precision	n	Mean	SD	%CV
Level 1	27	0.97	0.045	4.67%
Level 3	28	5.96	0.225	3.77%

Lactate


X: Abbott I-STAT System
Y: epoc System



n = 36
slope = 0.998
int't. = 0.113
Sy.x = 0.480
r = 0.996

Lactate


X: RAPIDLab 1265 Blood Gas System by Siemens Healthineers
Y: epoc System



n = 23
slope = 1.019
int't. = -0.207
Sy.x = 0.132
r = 0.999

Lactate


X: Radiometer ABL 700 System
Y: epoc System



n = 51
slope = 1.011
int't. = 0.101
Sy.x = 0.258
r = 0.995

Lactate


X: IL GEM PREMIER 4000 System
Y: epoc System



n = 46
slope = 1.025
int't. = 0.130
Sy.x = 0.564
r = 0.993

Lactate


X: Ortho Clinical Laboratories VITROS System
Y: epoc System



n = 42
slope = 0.938
int't. = 0.155
Sy.x = 0.398
r = 0.989

Lactate


X: Roche MODULAR System
Y: epoc System



n = 48
slope = 1.039
int't. = -0.067
Sy.x = 0.264
r = 0.996

Lactate

X: Dimension Integrated Chemistry System by Siemens Healthineers
Y: epoc System




n = 20
slope = 0.987
int't. = -0.033
Sy.x = 0.120
r = 0.999

BUN Method Comparison

BUN mg/dL				
Precision	n	Mean	SD	%CV
Level 1	137	49.9	1.12	2.2%
Level 3	132	4.9	0.13	2.7%

BUN


X: Dimension Integrated Chemistry System by Siemens Healthineers
Y: epoc System



n = 64
slope = 0.93
int't. = 0.3
Sy.x = 1.0
r = 0.999

BUN


X: Dimension Vista Intelligent Lab System by Siemens Healthineers
Y: epoc System



n = 146
slope = 0.95
int't. = 0.2
Sy.x = 1.6
r = 0.997

BUN


X: Beckman Coulter DXC System
Y: epoc System



n = 39
slope = 0.95
int't. = 1.3
Sy.x = 0.7
r = 0.999

BUN

X: Abbott ARCHITECT System
Y: epoc System




n = 86
slope = 0.93
int't. = 0.7
Sy.x = 0.7
r = 0.997

BUN

X: Roche COBAS System

Y: epoc System



$n = 121$

slope = 1.00

int't. = 0.2

$Sy.x = 1.8$


$r = 0.996$

Creatinine Method Comparison

Creatinine mg/dL				
Precision	n	Mean	SD	%CV
Level 1	20	0.91	0.045	4.93%
Level 3	20	4.54	0.191	4.21%

Creatinine


X: ADVIA Clinical Chemistry System by Siemens Healthineers
Y: epoc System



n = 53
slope = 1.063
int't. = -0.115
Sy.x = 0.207
r = 0.998

Creatinine


X: Abbott I-STAT System
Y: epoc System



n = 63
slope = 0.955
int't. = 0.075
Sy.x = 0.147
r = 0.999

Creatinine


X: Beckman Coulter AU680 System
Y: epoc System



n = 63
slope = 1.028
int't. = -0.008
Sy.x = 0.166
r = 0.999

Creatinine

X: Roche COBAS System
Y: epoc System



n = 50
slope = 1.069
int't. = -0.089
Sy.x = 0.201
r = 0.996

Notes

At Siemens Healthineers, our purpose is to enable healthcare providers to increase value by empowering them on their journey toward expanding precision medicine, transforming care delivery, and improving patient experience, all made possible by digitalizing healthcare.

An estimated 5 million patients globally benefit every day from our innovative technologies and services in the areas of diagnostic and therapeutic imaging, laboratory diagnostics, and molecular medicine, as well as digital health and enterprise services.

We are a leading medical technology company with over 120 years of experience and 18,000 patents globally. Through the dedication of more than 50,000 colleagues in 75 countries, we will continue to innovate and shape the future of healthcare.

ADVIA, Dimension, Dimension Vista, epoc, EXL, RAPIDLab, and all associated marks are trademarks of Siemens Healthcare Diagnostics Inc., or its affiliates. All other trademarks and brands are the property of their respective owners.

Product availability may vary from country to country and is subject to varying regulatory requirements. Please contact your local representative for availability.

Siemens Healthineers Headquarters

Siemens Healthcare GmbH
Henkestr. 127
91052 Erlangen, Germany
Phone: +49 9131 84-0
siemens-healthineers.com

Published by

Siemens Healthcare Diagnostics Inc.
Point of Care Diagnostics
2 Edgewater Drive
Norwood, MA 02062-4637
USA
Phone: +1 781-269-3000