

Release VA60A / 2026-07 / Revision 13427

syngo.share import

DICOM Conformance Statement

Overview

syngo.share import is able to load DICOM images, series and studies from specified directories or CD media and import them into syngo.share core and third-party archives.

Content and transfer

Table 1: Supported Storage SOP Classes

SOP Class	SOP Class UID	DIMSE services		DICOM web services		Media services		Function			
		SCU	SCP	UA	OS	FSC	FSR	Create	Display	Process	Archive
Stored Print Storage (Retired)	1.2.840.10008.5.1.1.27	Y	N	N	N	N	Y	N	N	N	N
Hardcopy Grayscale Image Storage (Retired)	1.2.840.10008.5.1.1.29	Y	N	N	N	N	Y	N	Y	N	N
Hardcopy Color Image Storage (Retired)	1.2.840.10008.5.1.1.30	Y	N	N	N	N	Y	N	N	N	N
Computed Radiography Image Storage	1.2.840.10008.5.1.4.1.1.1	Y	N	N	N	N	Y	N	Y	N	N
Digital X-Ray Image Storage - For Presentation	1.2.840.10008.5.1.4.1.1.1.1	Y	N	N	N	N	Y	N	Y	N	N
Digital X-Ray Image Storage - For Processing	1.2.840.10008.5.1.4.1.1.1.1.1	Y	N	N	N	N	Y	N	Y	N	N
Digital Mammography X-Ray Image Storage - For Presentation	1.2.840.10008.5.1.4.1.1.1.2	Y	N	N	N	N	Y	N	Y	N	N
Digital Mammography X-Ray Image Storage - For Processing	1.2.840.10008.5.1.4.1.1.1.2.1	Y	N	N	N	N	Y	N	Y	N	N
Digital Intra-Oral X-Ray Image Storage - For Presentation	1.2.840.10008.5.1.4.1.1.1.3	Y	N	N	N	N	Y	N	Y	N	N
Digital Intra-Oral X-Ray Image Storage - For Processing	1.2.840.10008.5.1.4.1.1.1.3.1	Y	N	N	N	N	Y	N	Y	N	N
CT Image Storage	1.2.840.10008.5.1.4.1.1.2	Y	N	N	N	N	Y	N	Y	N	N
Enhanced CT Image Storage	1.2.840.10008.5.1.4.1.1.2.1	Y	N	N	N	N	Y	N	Y	N	N
Legacy Converted Enhanced CT Image Storage	1.2.840.10008.5.1.4.1.1.2.2	Y	N	N	N	N	Y	N	Y	N	N
Ultrasound Multi-frame Image Storage (Retired)	1.2.840.10008.5.1.4.1.1.3	Y	N	N	N	N	Y	N	Y	N	N
Ultrasound Multi-frame Image Storage	1.2.840.10008.5.1.4.1.1.3.1	Y	N	N	N	N	Y	N	Y	N	N
MR Image Storage	1.2.840.10008.5.1.4.1.1.4	Y	N	N	N	N	Y	N	Y	N	N
Enhanced MR Image Storage	1.2.840.10008.5.1.4.1.1.4.1	Y	N	N	N	N	Y	N	Y	N	N

Table 1: Supported Storage SOP Classes ▼

▲ Table 1: Supported Storage SOP Classes ▼

SOP Class	SOP Class UID	DIMSE services		DICOM web services		Media services		Function			
		SCU	SCP	UA	OS	FSC	FSR	Create	Display	Process	Archive
MR Spectroscopy Storage	1.2.840.10008.5.1.4.1.1.4.2	Y	N	N	N	N	Y	N	N	N	N
Enhanced MR Color Image Storage	1.2.840.10008.5.1.4.1.1.4.3	Y	N	N	N	N	Y	N	Y	N	N
Legacy Converted Enhanced MR Image Storage	1.2.840.10008.5.1.4.1.1.4.4	Y	N	N	N	N	Y	N	N	N	N
Nuclear Medicine Image Storage (Retired)	1.2.840.10008.5.1.4.1.1.5	Y	N	N	N	N	Y	N	N	N	N
Ultrasound Image Storage (Retired)	1.2.840.10008.5.1.4.1.1.6	Y	N	N	N	N	Y	N	Y	N	N
Ultrasound Image Storage	1.2.840.10008.5.1.4.1.1.6.1	Y	N	N	N	N	Y	N	Y	N	N
Enhanced US Volume Storage	1.2.840.10008.5.1.4.1.1.6.2	Y	N	N	N	N	Y	N	Y	N	N
Photoacoustic Image Storage	1.2.840.10008.5.1.4.1.1.6.3	Y	N	N	N	N	Y	N	N	N	N
Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7	Y	N	N	N	N	Y	N	Y	N	N
Multi-frame Single Bit Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.1	Y	N	N	N	N	Y	S	Y	N	N
Multi-frame Grayscale Byte Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.2	Y	N	N	N	N	Y	S	Y	N	N
Multi-frame Grayscale Word Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.3	Y	N	N	N	N	Y	S	Y	N	N
Multi-frame True Color Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7.4	Y	N	N	N	N	Y	S	Y	N	N
Standalone Overlay Storage (Retired)	1.2.840.10008.5.1.4.1.1.8	Y	N	N	N	N	Y	N	N	N	N
Standalone Curve Storage (Retired)	1.2.840.10008.5.1.4.1.1.9	Y	N	N	N	N	Y	N	N	N	N
Waveform Storage - Trial (Retired)	1.2.840.10008.5.1.4.1.1.9.1	Y	N	N	N	N	Y	N	N	N	N
12-lead ECG Waveform Storage	1.2.840.10008.5.1.4.1.1.9.1.1	Y	N	N	N	N	Y	N	Y	N	N
General ECG Waveform Storage	1.2.840.10008.5.1.4.1.1.9.1.2	Y	N	N	N	N	Y	N	Y	N	N
Ambulatory ECG Waveform Storage	1.2.840.10008.5.1.4.1.1.9.1.3	Y	N	N	N	N	Y	N	N	N	N
General 32-bit ECG Waveform Storage	1.2.840.10008.5.1.4.1.1.9.1.4	Y	N	N	N	N	Y	N	N	N	N
Hemodynamic Waveform Storage	1.2.840.10008.5.1.4.1.1.9.2.1	Y	N	N	N	N	Y	N	Y	N	N
Cardiac Electrophysiology Waveform Storage	1.2.840.10008.5.1.4.1.1.9.3.1	Y	N	N	N	N	Y	N	N	N	N
Basic Voice Audio Waveform Storage	1.2.840.10008.5.1.4.1.1.9.4.1	Y	N	N	N	N	Y	N	N	N	N
General Audio Waveform Storage	1.2.840.10008.5.1.4.1.1.9.4.2	Y	N	N	N	N	Y	N	N	N	N
Arterial Pulse Waveform Storage	1.2.840.10008.5.1.4.1.1.9.5.1	Y	N	N	N	N	Y	N	N	N	N
Respiratory Waveform Storage	1.2.840.10008.5.1.4.1.1.9.6.1	Y	N	N	N	N	Y	N	N	N	N

▲ Table 1: Supported Storage SOP Classes ▼

^ Table 1: Supported Storage SOP Classes ▾

SOP Class	SOP Class UID	DIMSE services		DICOM web services		Media services		Function			
		SCU	SCP	UA	OS	FSC	FSR	Create	Display	Process	Archive
Multi-channel Respiratory Waveform Storage	1.2.840.10008.5.1.4.1.1.9.6.2	Y	N	N	N	N	Y	N	N	N	N
Routine Scalp Electroencephalogram Waveform Storage	1.2.840.10008.5.1.4.1.1.9.7.1	Y	N	N	N	N	Y	N	N	N	N
Electromyogram Waveform Storage	1.2.840.10008.5.1.4.1.1.9.7.2	Y	N	N	N	N	Y	N	N	N	N
Electrooculogram Waveform Storage	1.2.840.10008.5.1.4.1.1.9.7.3	Y	N	N	N	N	Y	N	N	N	N
Sleep Electroencephalogram Waveform Storage	1.2.840.10008.5.1.4.1.1.9.7.4	Y	N	N	N	N	Y	N	N	N	N
Body Position Waveform Storage	1.2.840.10008.5.1.4.1.1.9.8.1	Y	N	N	N	N	Y	N	N	N	N
Waveform Presentation State Storage	1.2.840.10008.5.1.4.1.1.9.100.1	Y	N	N	N	N	Y	N	N	N	N
Waveform Acquisition Presentation State Storage	1.2.840.10008.5.1.4.1.1.9.100.2	Y	N	N	N	N	Y	N	N	N	N
Standalone Modality LUT Storage (Retired)	1.2.840.10008.5.1.4.1.1.10	Y	N	N	N	N	Y	N	N	N	N
Standalone VOI LUT Storage (Retired)	1.2.840.10008.5.1.4.1.1.11	Y	N	N	N	N	Y	N	N	N	N
Grayscale Softcopy Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.1	Y	N	N	N	N	Y	N	N	N	N
Color Softcopy Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.2	Y	N	N	N	N	Y	N	N	N	N
Pseudo-Color Softcopy Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.3	Y	N	N	N	N	Y	N	N	N	N
Blending Softcopy Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.4	Y	N	N	N	N	Y	N	N	N	N
XA/XRF Grayscale Softcopy Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.5	Y	N	N	N	N	Y	N	N	N	N
Grayscale Planar MPR Volumetric Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.6	Y	N	N	N	N	Y	N	N	N	N
Compositing Planar MPR Volumetric Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.7	Y	N	N	N	N	Y	N	N	N	N
Advanced Blending Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.8	Y	N	N	N	N	Y	N	N	N	N
Volume Rendering Volumetric Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.9	Y	N	N	N	N	Y	N	N	N	N
Segmented Volume Rendering Volumetric Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.10	Y	N	N	N	N	Y	N	N	N	N
Multiple Volume Rendering Volumetric Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.11	Y	N	N	N	N	Y	N	N	N	N
Variable Modality LUT Softcopy Presentation State Storage	1.2.840.10008.5.1.4.1.1.11.12	Y	N	N	N	N	Y	N	N	N	N
X-Ray Angiographic Image Storage	1.2.840.10008.5.1.4.1.1.12.1	Y	N	N	N	N	Y	N	Y	N	N
Enhanced XA Image Storage	1.2.840.10008.5.1.4.1.1.12.1.1	Y	N	N	N	N	Y	N	Y	N	N
X-Ray Radiofluoroscopic Image Storage	1.2.840.10008.5.1.4.1.1.12.2	Y	N	N	N	N	Y	N	Y	N	N
Enhanced XRF Image Storage	1.2.840.10008.5.1.4.1.1.12.2.1	Y	N	N	N	N	Y	N	N	N	N

^ Table 1: Supported Storage SOP Classes ▾

Table 1: Supported Storage SOP Classes

SOP Class	SOP Class UID	DIMSE services		DICOM web services		Media services		Function			
		SCU	SCP	UA	OS	FSC	FSR	Create	Display	Process	Archive
X-Ray Angiographic Bi-Plane Image Storage (Retired)	1.2.840.10008.5.1.4.1.1.12.3	Y	N	N	N	N	Y	N	N	N	N
X-Ray 3D Angiographic Image Storage	1.2.840.10008.5.1.4.1.1.13.1.1	Y	N	N	N	N	Y	N	N	N	N
X-Ray 3D Craniofacial Image Storage	1.2.840.10008.5.1.4.1.1.13.1.2	Y	N	N	N	N	Y	N	N	N	N
Breast Tomosynthesis Image Storage	1.2.840.10008.5.1.4.1.1.13.1.3	Y	N	N	N	N	Y	N	Y	N	N
Breast Projection X-Ray Image Storage - For Presentation	1.2.840.10008.5.1.4.1.1.13.1.4	Y	N	N	N	N	Y	N	N	N	N
Breast Projection X-Ray Image Storage - For Processing	1.2.840.10008.5.1.4.1.1.13.1.5	Y	N	N	N	N	Y	N	N	N	N
Intravascular Optical Coherence Tomography Image Storage - For Presentation	1.2.840.10008.5.1.4.1.1.14.1	Y	N	N	N	N	Y	N	N	N	N
Intravascular Optical Coherence Tomography Image Storage - For Processing	1.2.840.10008.5.1.4.1.1.14.2	Y	N	N	N	N	Y	N	N	N	N
Nuclear Medicine Image Storage	1.2.840.10008.5.1.4.1.1.20	Y	N	N	N	N	Y	N	Y	N	N
Parametric Map Storage	1.2.840.10008.5.1.4.1.1.30	Y	N	N	N	N	Y	N	N	N	N
Raw Data Storage	1.2.840.10008.5.1.4.1.1.66	Y	N	N	N	N	Y	N	N	N	N
Spatial Registration Storage	1.2.840.10008.5.1.4.1.1.66.1	Y	N	N	N	N	Y	N	N	N	N
Spatial Fiducials Storage	1.2.840.10008.5.1.4.1.1.66.2	Y	N	N	N	N	Y	N	N	N	N
Deformable Spatial Registration Storage	1.2.840.10008.5.1.4.1.1.66.3	Y	N	N	N	N	Y	N	N	N	N
Segmentation Storage	1.2.840.10008.5.1.4.1.1.66.4	Y	N	N	N	N	Y	N	Y	N	N
Surface Segmentation Storage	1.2.840.10008.5.1.4.1.1.66.5	Y	N	N	N	N	Y	N	N	N	N
Tractography Results Storage	1.2.840.10008.5.1.4.1.1.66.6	Y	N	N	N	N	Y	N	N	N	N
Label Map Segmentation Storage	1.2.840.10008.5.1.4.1.1.66.7	Y	N	N	N	N	Y	N	N	N	N
Height Map Segmentation Storage	1.2.840.10008.5.1.4.1.1.66.8	Y	N	N	N	N	Y	N	N	N	N
Real World Value Mapping Storage	1.2.840.10008.5.1.4.1.1.67	Y	N	N	N	N	Y	N	N	N	N
Surface Scan Mesh Storage	1.2.840.10008.5.1.4.1.1.68.1	Y	N	N	N	N	Y	N	N	N	N
Surface Scan Point Cloud Storage	1.2.840.10008.5.1.4.1.1.68.2	Y	N	N	N	N	Y	N	N	N	N
VL Image Storage - Trial (Retired)	1.2.840.10008.5.1.4.1.1.77.1	Y	N	N	N	N	Y	N	N	N	N
VL Endoscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.1	Y	N	N	N	N	Y	N	Y	N	N
Video Endoscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.1.1	Y	N	N	N	N	Y	N	Y	N	N

Table 1: Supported Storage SOP Classes

Table 1: Supported Storage SOP Classes

SOP Class	SOP Class UID	DIMSE services		DICOM web services		Media services		Function			
		SCU	SCP	UA	OS	FSC	FSR	Create	Display	Process	Archive
VL Microscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.2	Y	N	N	N	N	Y	N	Y	N	N
Video Microscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.2.1	Y	N	N	N	N	Y	N	Y	N	N
VL Slide-Coordinates Microscopic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.3	Y	N	N	N	N	Y	N	N	N	N
VL Photographic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.4	Y	N	N	N	N	Y	N	Y	N	N
Video Photographic Image Storage	1.2.840.10008.5.1.4.1.1.77.1.4.1	Y	N	N	N	N	Y	N	Y	N	N
Ophthalmic Photography 8 Bit Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.1	Y	N	N	N	N	Y	N	Y	N	N
Ophthalmic Photography 16 Bit Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.2	Y	N	N	N	N	Y	N	N	N	N
Stereometric Relationship Storage	1.2.840.10008.5.1.4.1.1.77.1.5.3	Y	N	N	N	N	Y	N	N	N	N
Ophthalmic Tomography Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.4	Y	N	N	N	N	Y	N	Y	N	N
Wide Field Ophthalmic Photography Stereographic Projection Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.5	Y	N	N	N	N	Y	N	N	N	N
Wide Field Ophthalmic Photography 3D Coordinates Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.6	Y	N	N	N	N	Y	N	N	N	N
Ophthalmic Optical Coherence Tomography En Face Image Storage	1.2.840.10008.5.1.4.1.1.77.1.5.7	Y	N	N	N	N	Y	N	N	N	N
Ophthalmic Optical Coherence Tomography B-scan Volume Analysis Storage	1.2.840.10008.5.1.4.1.1.77.1.5.8	Y	N	N	N	N	Y	N	N	N	N
VL Whole Slide Microscopy Image Storage	1.2.840.10008.5.1.4.1.1.77.1.6	Y	N	N	N	N	Y	N	N	N	N
Dermoscopic Photography Image Storage	1.2.840.10008.5.1.4.1.1.77.1.7	Y	N	N	N	N	Y	N	N	N	N
Confocal Microscopy Image Storage	1.2.840.10008.5.1.4.1.1.77.1.8	Y	N	N	N	N	Y	N	N	N	N
Confocal Microscopy Tiled Pyramidal Image Storage	1.2.840.10008.5.1.4.1.1.77.1.9	Y	N	N	N	N	Y	N	N	N	N
VL Multi-frame Image Storage - Trial (Retired)	1.2.840.10008.5.1.4.1.1.77.2	Y	N	N	N	N	Y	N	N	N	N
Lensometry Measurements Storage	1.2.840.10008.5.1.4.1.1.78.1	Y	N	N	N	N	Y	N	N	N	N
Autorefracton Measurements Storage	1.2.840.10008.5.1.4.1.1.78.2	Y	N	N	N	N	Y	N	N	N	N
Keratometry Measurements Storage	1.2.840.10008.5.1.4.1.1.78.3	Y	N	N	N	N	Y	N	N	N	N
Subjective Refraction Measurements Storage	1.2.840.10008.5.1.4.1.1.78.4	Y	N	N	N	N	Y	N	N	N	N
Visual Acuity Measurements Storage	1.2.840.10008.5.1.4.1.1.78.5	Y	N	N	N	N	Y	N	N	N	N
Spectacle Prescription Report Storage	1.2.840.10008.5.1.4.1.1.78.6	Y	N	N	N	N	Y	N	N	N	N
Ophthalmic Axial Measurements Storage	1.2.840.10008.5.1.4.1.1.78.7	Y	N	N	N	N	Y	N	N	N	N

Table 1: Supported Storage SOP Classes

▲ Table 1: Supported Storage SOP Classes ▼

SOP Class	SOP Class UID	DIMSE services		DICOM web services		Media services		Function			
		SCU	SCP	UA	OS	FSC	FSR	Create	Display	Process	Archive
Intraocular Lens Calculations Storage	1.2.840.10008.5.1.4.1.1.78.8	Y	N	N	N	N	Y	N	N	N	N
Macular Grid Thickness and Volume Report Storage	1.2.840.10008.5.1.4.1.1.79.1	Y	N	N	N	N	Y	N	N	N	N
Ophthalmic Visual Field Static Perimetry Measurements Storage	1.2.840.10008.5.1.4.1.1.80.1	Y	N	N	N	N	Y	N	N	N	N
Ophthalmic Thickness Map Storage	1.2.840.10008.5.1.4.1.1.81.1	Y	N	N	N	N	Y	N	N	N	N
Corneal Topography Map Storage	1.2.840.10008.5.1.4.1.1.82.1	Y	N	N	N	N	Y	N	N	N	N
Text SR Storage - Trial (Retired)	1.2.840.10008.5.1.4.1.1.88.1	Y	N	N	N	N	Y	N	N	N	N
Audio SR Storage - Trial (Retired)	1.2.840.10008.5.1.4.1.1.88.2	Y	N	N	N	N	Y	N	N	N	N
Detail SR Storage - Trial (Retired)	1.2.840.10008.5.1.4.1.1.88.3	Y	N	N	N	N	Y	N	N	N	N
Comprehensive SR Storage - Trial (Retired)	1.2.840.10008.5.1.4.1.1.88.4	Y	N	N	N	N	Y	N	N	N	N
Basic Text SR Storage	1.2.840.10008.5.1.4.1.1.88.11	Y	N	N	N	N	Y	N	Y	N	N
Enhanced SR Storage	1.2.840.10008.5.1.4.1.1.88.22	Y	N	N	N	N	Y	N	Y	N	N
Comprehensive SR Storage	1.2.840.10008.5.1.4.1.1.88.33	Y	N	N	N	N	Y	N	Y	N	N
Comprehensive 3D SR Storage	1.2.840.10008.5.1.4.1.1.88.34	Y	N	N	N	N	Y	N	N	N	N
Extensible SR Storage	1.2.840.10008.5.1.4.1.1.88.35	Y	N	N	N	N	Y	N	N	N	N
Procedure Log Storage	1.2.840.10008.5.1.4.1.1.88.40	Y	N	N	N	N	Y	N	N	N	N
Mammography CAD SR Storage	1.2.840.10008.5.1.4.1.1.88.50	Y	N	N	N	N	Y	N	Y	N	N
Key Object Selection Document Storage	1.2.840.10008.5.1.4.1.1.88.59	Y	N	N	N	N	Y	N	Y	N	N
Chest CAD SR Storage	1.2.840.10008.5.1.4.1.1.88.65	Y	N	N	N	N	Y	N	N	N	N
X-Ray Radiation Dose SR Storage	1.2.840.10008.5.1.4.1.1.88.67	Y	N	N	N	N	Y	N	Y	N	N
Radiopharmaceutical Radiation Dose SR Storage	1.2.840.10008.5.1.4.1.1.88.68	Y	N	N	N	N	Y	N	N	N	N
Colon CAD SR Storage	1.2.840.10008.5.1.4.1.1.88.69	Y	N	N	N	N	Y	N	N	N	N
Implantation Plan SR Storage	1.2.840.10008.5.1.4.1.1.88.70	Y	N	N	N	N	Y	N	N	N	N
Acquisition Context SR Storage	1.2.840.10008.5.1.4.1.1.88.71	Y	N	N	N	N	Y	N	N	N	N
Simplified Adult Echo SR Storage	1.2.840.10008.5.1.4.1.1.88.72	Y	N	N	N	N	Y	N	N	N	N
Patient Radiation Dose SR Storage	1.2.840.10008.5.1.4.1.1.88.73	Y	N	N	N	N	Y	N	N	N	N
Planned Imaging Agent Administration SR Storage	1.2.840.10008.5.1.4.1.1.88.74	Y	N	N	N	N	Y	N	N	N	N

▲ Table 1: Supported Storage SOP Classes ▼

▲ Table 1: Supported Storage SOP Classes ▼

SOP Class	SOP Class UID	DIMSE services		DICOM web services		Media services		Function			
		SCU	SCP	UA	OS	FSC	FSR	Create	Display	Process	Archive
Performed Imaging Agent Administration SR Storage	1.2.840.10008.5.1.4.1.1.88.75	Y	N	N	N	N	Y	N	N	N	N
Enhanced X-Ray Radiation Dose SR Storage	1.2.840.10008.5.1.4.1.1.88.76	Y	N	N	N	N	Y	N	N	N	N
Waveform Annotation SR Storage	1.2.840.10008.5.1.4.1.1.88.77	Y	N	N	N	N	Y	N	N	N	N
Content Assessment Results Storage	1.2.840.10008.5.1.4.1.1.90.1	Y	N	N	N	N	Y	N	N	N	N
Microscopy Bulk Simple Annotations Storage	1.2.840.10008.5.1.4.1.1.91.1	Y	N	N	N	N	Y	N	N	N	N
Encapsulated PDF Storage	1.2.840.10008.5.1.4.1.1.104.1	Y	N	N	N	N	Y	S	Y	N	N
Encapsulated CDA Storage	1.2.840.10008.5.1.4.1.1.104.2	Y	N	N	N	N	Y	N	N	N	N
Encapsulated STL Storage	1.2.840.10008.5.1.4.1.1.104.3	Y	N	N	N	N	Y	N	N	N	N
Encapsulated OBJ Storage	1.2.840.10008.5.1.4.1.1.104.4	Y	N	N	N	N	Y	N	N	N	N
Encapsulated MTL Storage	1.2.840.10008.5.1.4.1.1.104.5	Y	N	N	N	N	Y	N	N	N	N
Positron Emission Tomography Image Storage	1.2.840.10008.5.1.4.1.1.128	Y	N	N	N	N	Y	N	Y	N	N
Legacy Converted Enhanced PET Image Storage	1.2.840.10008.5.1.4.1.1.128.1	Y	N	N	N	N	Y	N	N	N	N
Standalone PET Curve Storage (Retired)	1.2.840.10008.5.1.4.1.1.129	Y	N	N	N	N	Y	N	N	N	N
Enhanced PET Image Storage	1.2.840.10008.5.1.4.1.1.130	Y	N	N	N	N	Y	N	N	N	N
Basic Structured Display Storage	1.2.840.10008.5.1.4.1.1.131	Y	N	N	N	N	Y	N	N	N	N
CT Performed Procedure Protocol Storage	1.2.840.10008.5.1.4.1.1.200.2	Y	N	N	N	N	Y	N	N	N	N
XA Performed Procedure Protocol Storage	1.2.840.10008.5.1.4.1.1.200.8	Y	N	N	N	N	Y	N	N	N	N
RT Image Storage	1.2.840.10008.5.1.4.1.1.481.1	Y	N	N	N	N	Y	N	Y	N	N
RT Dose Storage	1.2.840.10008.5.1.4.1.1.481.2	Y	N	N	N	N	Y	N	Y	N	N
RT Structure Set Storage	1.2.840.10008.5.1.4.1.1.481.3	Y	N	N	N	N	Y	N	N	N	N
RT Beams Treatment Record Storage	1.2.840.10008.5.1.4.1.1.481.4	Y	N	N	N	N	Y	N	N	N	N
RT Plan Storage	1.2.840.10008.5.1.4.1.1.481.5	Y	N	N	N	N	Y	N	N	N	N
RT Brachy Treatment Record Storage	1.2.840.10008.5.1.4.1.1.481.6	Y	N	N	N	N	Y	N	N	N	N
RT Treatment Summary Record Storage	1.2.840.10008.5.1.4.1.1.481.7	Y	N	N	N	N	Y	N	N	N	N
RT Ion Plan Storage	1.2.840.10008.5.1.4.1.1.481.8	Y	N	N	N	N	Y	N	N	N	N
RT Ion Beams Treatment Record Storage	1.2.840.10008.5.1.4.1.1.481.9	Y	N	N	N	N	Y	N	N	N	N

▲ Table 1: Supported Storage SOP Classes ▼

▲ Table 1: Supported Storage SOP Classes ▼

SOP Class	SOP Class UID	DIMSE services		DICOM web services		Media services		Function			
		SCU	SCP	UA	OS	FSC	FSR	Create	Display	Process	Archive
RT Physician Intent Storage	1.2.840.10008.5.1.4.1.1.481.10	Y	N	N	N	N	Y	N	N	N	N
RT Segment Annotation Storage	1.2.840.10008.5.1.4.1.1.481.11	Y	N	N	N	N	Y	N	N	N	N
RT Radiation Set Storage	1.2.840.10008.5.1.4.1.1.481.12	Y	N	N	N	N	Y	N	N	N	N
C-Arm Photon-Electron Radiation Storage	1.2.840.10008.5.1.4.1.1.481.13	Y	N	N	N	N	Y	N	N	N	N
Tomotherapeutic Radiation Storage	1.2.840.10008.5.1.4.1.1.481.14	Y	N	N	N	N	Y	N	N	N	N
Robotic-Arm Radiation Storage	1.2.840.10008.5.1.4.1.1.481.15	Y	N	N	N	N	Y	N	N	N	N
RT Radiation Record Set Storage	1.2.840.10008.5.1.4.1.1.481.16	Y	N	N	N	N	Y	N	N	N	N
RT Radiation Salvage Record Storage	1.2.840.10008.5.1.4.1.1.481.17	Y	N	N	N	N	Y	N	N	N	N
Tomotherapeutic Radiation Record Storage	1.2.840.10008.5.1.4.1.1.481.18	Y	N	N	N	N	Y	N	N	N	N
C-Arm Photon-Electron Radiation Record Storage	1.2.840.10008.5.1.4.1.1.481.19	Y	N	N	N	N	Y	N	N	N	N
Robotic Radiation Record Storage	1.2.840.10008.5.1.4.1.1.481.20	Y	N	N	N	N	Y	N	N	N	N
RT Radiation Set Delivery Instruction Storage	1.2.840.10008.5.1.4.1.1.481.21	Y	N	N	N	N	Y	N	N	N	N
RT Treatment Preparation Storage	1.2.840.10008.5.1.4.1.1.481.22	Y	N	N	N	N	Y	N	N	N	N
Enhanced RT Image Storage	1.2.840.10008.5.1.4.1.1.481.23	Y	N	N	N	N	Y	N	N	N	N
Enhanced Continuous RT Image Storage	1.2.840.10008.5.1.4.1.1.481.24	Y	N	N	N	N	Y	N	N	N	N
RT Patient Position Acquisition Instruction Storage	1.2.840.10008.5.1.4.1.1.481.25	Y	N	N	N	N	Y	N	N	N	N
DICOS CT Image Storage	1.2.840.10008.5.1.4.1.1.501.1	Y	N	N	N	N	Y	N	N	N	N
DICOS Digital X-Ray Image Storage - For Presentation	1.2.840.10008.5.1.4.1.1.501.2.1	Y	N	N	N	N	Y	N	Y	N	N
DICOS Digital X-Ray Image Storage - For Processing	1.2.840.10008.5.1.4.1.1.501.2.2	Y	N	N	N	N	Y	N	N	N	N
DICOS Threat Detection Report Storage	1.2.840.10008.5.1.4.1.1.501.3	Y	N	N	N	N	Y	N	N	N	N
DICOS 2D AIT Storage	1.2.840.10008.5.1.4.1.1.501.4	Y	N	N	N	N	Y	N	N	N	N
DICOS 3D AIT Storage	1.2.840.10008.5.1.4.1.1.501.5	Y	N	N	N	N	Y	N	N	N	N
DICOS Quadrupole Resonance (QR) Storage	1.2.840.10008.5.1.4.1.1.501.6	Y	N	N	N	N	Y	N	N	N	N
Eddy Current Image Storage	1.2.840.10008.5.1.4.1.1.601.1	Y	N	N	N	N	Y	N	N	N	N
Eddy Current Multi-frame Image Storage	1.2.840.10008.5.1.4.1.1.601.2	Y	N	N	N	N	Y	N	N	N	N
Thermography Image Storage	1.2.840.10008.5.1.4.1.1.601.3	Y	N	N	N	N	Y	N	N	N	N

▲ Table 1: Supported Storage SOP Classes ▼

▲ Table 1: Supported Storage SOP Classes

SOP Class	SOP Class UID	DIMSE services		DICOM web services		Media services		Function			
		SCU	SCP	UA	OS	FSC	FSR	Create	Display	Process	Archive
Thermography Multi-frame Image Storage	1.2.840.10008.5.1.4.1.1.601.4	Y	N	N	N	N	Y	N	N	N	N
Ultrasound Waveform Storage	1.2.840.10008.5.1.4.1.1.601.5	Y	N	N	N	N	Y	N	N	N	N
RT Beams Delivery Instruction Storage - Trial (Retired)	1.2.840.10008.5.1.4.34.1	Y	N	N	N	N	Y	N	N	N	N
RT Beams Delivery Instruction Storage	1.2.840.10008.5.1.4.34.7	Y	N	N	N	N	Y	N	N	N	N
RT Brachy Application Setup Delivery Instruction Storage	1.2.840.10008.5.1.4.34.10	Y	N	N	N	N	Y	N	N	N	N

Table 2: Supported Storage Transfer Syntaxes

Transfer Syntax name	Transfer Syntax UID
Implicit VR Little Endian: Default Transfer Syntax for DICOM	1.2.840.10008.1.2
Explicit VR Little Endian	1.2.840.10008.1.2.1
Encapsulated Uncompressed Explicit VR Little Endian	1.2.840.10008.1.2.1.98
Deflated Explicit VR Little Endian	1.2.840.10008.1.2.1.99
Explicit VR Big Endian (Retired)	1.2.840.10008.1.2.2
JPEG Baseline (Process 1): Default Transfer Syntax for Lossy JPEG 8 Bit Image Compression	1.2.840.10008.1.2.4.50
JPEG Extended (Process 2 & 4): Default Transfer Syntax for Lossy JPEG 12 Bit Image Compression (Process 4 only)	1.2.840.10008.1.2.4.51
JPEG Lossless, Non-Hierarchical (Process 14)	1.2.840.10008.1.2.4.57
JPEG Lossless, Non-Hierarchical, First-Order Prediction (Process 14 [Selection Value 1]): Default Transfer Syntax for Lossless JPEG Image Compression	1.2.840.10008.1.2.4.70
JPEG-LS Lossless Image Compression	1.2.840.10008.1.2.4.80
JPEG-LS Lossy (Near-Lossless) Image Compression	1.2.840.10008.1.2.4.81
JPEG 2000 Image Compression (Lossless Only)	1.2.840.10008.1.2.4.90
JPEG 2000 Image Compression	1.2.840.10008.1.2.4.91
JPEG 2000 Part 2 Multi-component Image Compression (Lossless Only)	1.2.840.10008.1.2.4.92
JPEG 2000 Part 2 Multi-component Image Compression	1.2.840.10008.1.2.4.93
MPEG2 Main Profile / Main Level	1.2.840.10008.1.2.4.100

Table 2: Supported Storage Transfer Syntaxes ▼

▲ Table 2: Supported Storage Transfer Syntaxes

Transfer Syntax name	Transfer Syntax UID
Fragmentable MPEG2 Main Profile / Main Level	1.2.840.10008.1.2.4.100.1
MPEG2 Main Profile / High Level	1.2.840.10008.1.2.4.101
Fragmentable MPEG2 Main Profile / High Level	1.2.840.10008.1.2.4.101.1
MPEG-4 AVC/H.264 High Profile / Level 4.1	1.2.840.10008.1.2.4.102
Fragmentable MPEG-4 AVC/H.264 High Profile / Level 4.1	1.2.840.10008.1.2.4.102.1
MPEG-4 AVC/H.264 BD-compatible High Profile / Level 4.1	1.2.840.10008.1.2.4.103
Fragmentable MPEG-4 AVC/H.264 BD-compatible High Profile / Level 4.1	1.2.840.10008.1.2.4.103.1
MPEG-4 AVC/H.264 High Profile / Level 4.2 For 2D Video	1.2.840.10008.1.2.4.104
Fragmentable MPEG-4 AVC/H.264 High Profile / Level 4.2 For 2D Video	1.2.840.10008.1.2.4.104.1
MPEG-4 AVC/H.264 High Profile / Level 4.2 For 3D Video	1.2.840.10008.1.2.4.105
Fragmentable MPEG-4 AVC/H.264 High Profile / Level 4.2 For 3D Video	1.2.840.10008.1.2.4.105.1
MPEG-4 AVC/H.264 Stereo High Profile / Level 4.2	1.2.840.10008.1.2.4.106
Fragmentable MPEG-4 AVC/H.264 Stereo High Profile / Level 4.2	1.2.840.10008.1.2.4.106.1
HEVC/H.265 Main Profile / Level 5.1	1.2.840.10008.1.2.4.107
HEVC/H.265 Main 10 Profile / Level 5.1	1.2.840.10008.1.2.4.108
JPEG XL Lossless	1.2.840.10008.1.2.4.110
JPEG XL JPEG Recompression	1.2.840.10008.1.2.4.111
JPEG XL	1.2.840.10008.1.2.4.112
High-Throughput JPEG 2000 Image Compression (Lossless Only)	1.2.840.10008.1.2.4.201
High-Throughput JPEG 2000 with RPCL Options Image Compression (Lossless Only)	1.2.840.10008.1.2.4.202
High-Throughput JPEG 2000 Image Compression	1.2.840.10008.1.2.4.203
RLE Lossless	1.2.840.10008.1.2.5
Deflated Image Frame Compression	1.2.840.10008.1.2.8.1

DIMSE services

Verification

Table 3: Supported Verification SOP Classes

SOP Class	SOP Class UID	Transfer Syntax	Transfer Syntax UID	SCU	SCP
Verification	1.2.840.10008.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	Y	N
		Explicit VR Little Endian	1.2.840.10008.1.2.1	Y	N
		Explicit VR Big Endian (Retired)	1.2.840.10008.1.2.2	Y	N

Workflow management

Table 4: Supported Storage Commitment SOP Classes

SOP Class	SOP Class UID	Transfer Syntax	Transfer Syntax UID	SCU	SCP
Storage Commitment Push Model	1.2.840.10008.1.20.1	Implicit VR Little Endian	1.2.840.10008.1.2	Y	N
		Explicit VR Little Endian	1.2.840.10008.1.2.1	Y	N
		Explicit VR Big Endian (Retired)	1.2.840.10008.1.2.2	Y	N

Table 5: Supported Modality Performed Procedure Step SOP Classes

SOP Class	SOP Class UID	Transfer Syntax	Transfer Syntax UID	SCU	SCP
Modality Performed Procedure Step	1.2.840.10008.3.1.2.3.3	Implicit VR Little Endian	1.2.840.10008.1.2	Y	N
		Explicit VR Little Endian	1.2.840.10008.1.2.1	Y	N
		Explicit VR Big Endian (Retired)	1.2.840.10008.1.2.2	Y	N

Query/Retrieve

Table 6: Supported Query/Retrieve SOP Classes

SOP Class	SOP Class UID	Transfer Syntax	Transfer Syntax UID	SCU	SCP
Patient Root Query/Retrieve Information Model – FIND	1.2.840.10008.5.1.4.1.2.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	Y	N
		Explicit VR Little Endian	1.2.840.10008.1.2.1	Y	N
		Explicit VR Big Endian (Retired)	1.2.840.10008.1.2.2	Y	N
Study Root Query/Retrieve Information Model – FIND	1.2.840.10008.5.1.4.1.2.2.1	Implicit VR Little Endian	1.2.840.10008.1.2	Y	N
		Explicit VR Little Endian	1.2.840.10008.1.2.1	Y	N
		Explicit VR Big Endian (Retired)	1.2.840.10008.1.2.2	Y	N
Study Root Query/Retrieve Information Model – MOVE	1.2.840.10008.5.1.4.1.2.2.2	Implicit VR Little Endian	1.2.840.10008.1.2	Y	N
		Explicit VR Little Endian	1.2.840.10008.1.2.1	Y	N
		Explicit VR Big Endian (Retired)	1.2.840.10008.1.2.2	Y	N
Modality Worklist Information Model – FIND	1.2.840.10008.5.1.4.31	Implicit VR Little Endian	1.2.840.10008.1.2	Y	N
		Explicit VR Little Endian	1.2.840.10008.1.2.1	Y	N
		Explicit VR Big Endian (Retired)	1.2.840.10008.1.2.2	Y	N

Contents

Overview	2
Content and transfer	2
DIMSE services	11
Verification	12
Workflow management	12
Query/Retrieve	13
1 Introduction	16
1.1 Remarks	16
1.2 Abbreviations	16
2 Implementation model	17
2.1 Application data flow diagram	17
2.2 Functional definition of AEs	17
2.3 Sequencing of real-world activities	17
3 Application Entity specifications	18
3.1 Supported SOP Classes and Transfer Syntaxes	18
3.2 Association establishment policies	18
3.2.1 General	18
3.2.2 Number of associations	18
3.2.3 Asynchronous nature	18
3.2.4 Implementation identifying information	18
3.2.5 Association initiation policy by real-world activity	18
4 DICOM media AE specification	19
4.1 Implementation model	19
4.1.1 Application data flow diagram	19
4.1.2 Functional definitions of AEs	19
4.1.3 Sequencing of real-world activities	19
4.1.4 File meta information	19
4.2 Application Entity specification	19
4.2.1 Real-world activities	19
4.3 Augmented and private application profiles	20
5 Configuration	21
6 Support of extended character sets	22
6.1 Supported character sets	22
6.2 Usage of Specific Character Set in C-FIND and C-MOVE requests	22
A DICOM element list for Query/Retrieve Service Classes	23
A.1 Supported C-FIND element requests on the patient level	23
A.2 Supported C-FIND element requests on the study level	23
A.3 Supported C-FIND element requests on the series level	24
B List of elements for modality worklist C-FIND requests	25

List of figures

1	<i>syngo.share</i> import data flow diagram	17
---	---	----

List of tables

1	Supported Storage SOP Classes	2
2	Supported Storage Transfer Syntaxes	10
3	Supported Verification SOP Classes	12
4	Supported Storage Commitment SOP Classes	12
5	Supported Modality Performed Procedure Step SOP Classes	12
6	Supported Query/Retrieve SOP Classes	13
7	<i>syngo.share</i> import AE-related application profiles, real-world activity roles, and roles for interchanging	19
8	Supported character sets	22
9	Supported C-FIND elements on the patient level	23
10	Supported C-FIND elements on the study level	23
11	Supported C-FIND elements on the series level	24
12	Supported C-FIND modality worklist keywords	25

1 Introduction

This document is a DICOM Conformance Statement that describes the DICOM capabilities of *syngo.share import*.

1.1 Remarks

This Conformance Statement is intended to aid in the validation of the integration of *syngo.share import* within a DICOM environment. This statement is not meant to replace the validation with other DICOM equipment to ensure the intended, proper exchange of information. Thus, it is still important to ensure the proper interoperability of the intended DICOM integration.

The user must be aware of the following issues:

- The comparison of different Conformance Statements should be the first step towards an assessment of the interoperability within a DICOM environment.
- Testing procedures should be defined to validate the desired level of connectivity.

1.2 Abbreviations


AE	Application Entity
AET	Application Entity Title
CAD	Computer Aided Detection
CDA	Clinical Document Architecture
CID	Context Identifier
DCS	DICOM Conformance Statement
DICOM	Digital Imaging and Communications in Medicine
FSC	File-Set Creator
FSR	File-Set Reader
IHE	Integrating the Healthcare Enterprise
IOD	Information Object Definition
ISO	International Organization for Standardization
MPPS	Modality Performed Procedure Step
OS	Origin Server
PDU	Protocol Data Unit
QIDO-RS	Query based on ID for DICOM Objects by RESTful Services
SCP	Service Class Provider
SCU	Service Class User
SOP	Service-Object Pair
SR	Structured Reporting
TCP/IP	Transmission Control Protocol/Internet Protocol
UA	User Agent
UID	Unique Identifier
VR	Value Representation
WADO-RS	Web Access to DICOM Objects by RESTful Services

2 Implementation model

2.1 Application data flow diagram

syngo.share import (i.e. the FSR) is able to read DICOM files from portable media such as CDs, from network directories or from the local file system. Loaded DICOM file sets can be imported directly into *syngo.share core*. Within the context of this Conformance Statement, the importing functionality of *syngo.share import* is referred to as the *syngo.share import* FSR AE.

Figure 1: *syngo.share import* data flow diagram



2.2 Functional definition of AEs

The *syngo.share import* FSR AE is able to read user-selected DICOM files like DICOMDIRs or image objects which are compliant to DICOM PS3.10. These files can be located either on the local file system or on DICOM PS3.12-compliant media.

2.3 Sequencing of real-world activities

- The user can request the services of the *syngo.share import* AE at any time through the user interface.

3 Application Entity specifications

The *syngo.share* import FSR solely provides functionalities for handling DICOM media. The DICOM Storage SOP Classes listed in Table 1 are supported for storage.

3.1 Supported SOP Classes and Transfer Syntaxes

The supported DICOM SOP Classes are described in Table 1. Table 2 lists the supported C-STORE Transfer Syntaxes.

3.2 Association establishment policies

3.2.1 General

The *syngo.share* import AE supports TCP/IP. When a user requests a C-STORE operation, it attempts to establish an association with a remote AE. The host, port and remote AE title are defined within the client configuration.

3.2.2 Number of associations

The *syngo.share* import AE establishes one associations per storage task.

3.2.3 Asynchronous nature

The *syngo.share* import AE only allows a single outstanding operation on each association, meaning that it does not perform asynchronous negotiation.

3.2.4 Implementation identifying information

- **Implementation Class UID**
1.2.276.0.7230010.3.0.3.6.1
- **Implementation Version Name**
OFFIS_DCMTK_361

3.2.5 Association initiation policy by real-world activity

The *syngo.share* import AE initiates an association with a remote AE for C-STORE requests. The proposed Transfer Syntaxes for association establishment for C-STORE are defined in Table 2.

4 DICOM media AE specification

This chapter describes the DICOM media functionalities of the *syngo.share* import AE.

4.1 Implementation model

4.1.1 Application data flow diagram

See [Section 2.1](#) for the data flow diagram.

4.1.2 Functional definitions of AEs

The *syngo.share* import AE implements standard DICOM-conformant Service Classes for the reading of DICOM file sets (according to DICOM PS3.10). At least, General Purpose CD-R Interchange Profiles are supported.

4.1.3 Sequencing of real-world activities

The DICOM media functionalities of the *syngo.share* import AE can be used at any time through their user interfaces.

4.1.4 File meta information

- **Implementation Class UID**
1.2.276.0.7230010.3.0.3.6.1
- **Implementation Version Name**
OFFIS_DCMTK_361

4.2 Application Entity specification

See [Table 1](#) for supported SOP Classes for the import of media.

Table 7: *syngo.share* import AE-related application profiles, real-world activity roles, and roles for interchanging

Supported application profiles	Real-world activity	Role	SC option
<i>syngo.share</i> import FSR	Read DICOM file set	FSR	Interchange

4.2.1 Real-world activities

4.2.1.1 Import Media

The user can choose whether a complete file set or just parts of it are read for import.

4.2.1.2 Reading DICOMDIR keys

All mandatory DICOMDIR keys are required in order to correctly structure the images within the file sets.

4.3 Augmented and private application profiles

Not used.

5 Configuration

`syngo.share import` provides user interfaces in order to facilitate configuration.

6 Support of extended character sets

6.1 Supported character sets

Table 8 contains the character sets which are supported with and without code extension techniques.

If the given specific character set does not correspond to the characters in an IOD or the specific character set is invalid, the DICOM dataset is not processed. In these cases, a configuration can be used to correct the character set. Due to practical reasons, the specific character set is used instead of the default character repertoire to process tags with VR CS.

Table 8: Supported character sets

MIME name	Without code extensions	With code extensions
US-ASCII		ISO 2022 IR 6
ISO-8859-1	ISO_IR 100	ISO 2022 IR 100
ISO-8859-2	ISO_IR 101	ISO 2022 IR 101
ISO-8859-3	ISO_IR 109	ISO 2022 IR 109
ISO-8859-4	ISO_IR 110	ISO 2022 IR 110
ISO-8859-5	ISO_IR 144	ISO 2022 IR 144
ISO-8859-6	ISO_IR 127	ISO 2022 IR 127
ISO-8859-7	ISO_IR 126	ISO 2022 IR 126
ISO-8859-8	ISO_IR 138	ISO 2022 IR 138
ISO-8859-9	ISO_IR 148	ISO 2022 IR 148
JIS_X0201	ISO_IR 13	ISO 2022 IR 13
TIS-620	ISO_IR 166	ISO 2022 IR 166
JIS_X0208-1990	-	ISO 2022 IR 87
JIS_X0212-1990	-	ISO 2022 IR 159
KS_X_1001-1997	-	ISO 2022 IR 149
GB2312	-	ISO 2022 IR 58
UTF-8	ISO_IR 192	-
GB18030	GB18030	-
GBK	GBK	-

6.2 Usage of Specific Character Set in C-FIND and C-MOVE requests

syngo.share import AE as C-FIND and C-MOVE SCU uses Specific Character Set (0008,0005) with a value of ISO_IR 192 for requests.

A DICOM element list for Query/Retrieve Service Classes

The following table lists the DICOM keys used for matching or query on the according level in C-FIND requests.

A.1 Supported C-FIND element requests on the patient level

Table 9: Supported C-FIND elements on the patient level

Keyword	Tag	Used for matching
PatientName	(0010,0010)	Y
PatientID	(0010,0020)	Y
IssuerOfPatientIDQualifiersSequence	(0010,0024)	Y
PatientBirthDate	(0010,0030)	Y
PatientBirthTime	(0010,0032)	Y
PatientSex	(0010,0040)	Y
OtherPatientIDs	(0010,1000)	Y
PatientBirthName	(0010,1005)	Y

A.2 Supported C-FIND element requests on the study level

In case of a Study Root Query/Retrieve where no patient level exists, all mentioned patient level keys are supported on the study level.

Table 10: Supported C-FIND elements on the study level

Keyword	Tag	Used for matching
StudyDate	(0008,0020)	Y
StudyTime	(0008,0030)	Y
AccessionNumber	(0008,0050)	Y
ModalitiesInStudy	(0008,0061)	Y
StudyDescription	(0008,1030)	Y
StudyInstanceUID	(0020,000D)	Y
NumberOfStudyRelatedSeries	(0020,1206)	N
NumberOfStudyRelatedInstances	(0020,1208)	N

A.3 Supported C-FIND element requests on the series level

Table 11: Supported C-FIND elements on the series level

Keyword	Tag	Used for matching
SeriesDate	(0008,0021)	N
SeriesTime	(0008,0031)	N
Modality	(0008,0060)	N
SeriesDescription	(0008,103E)	N
SeriesInstanceUID	(0020,000E)	Y
NumberOfSeriesRelatedInstances	(0020,1209)	N

B List of elements for modality worklist C-FIND requests

Table 12: Supported C-FIND modality worklist keywords

Key	Tag	Used for matching
AccessionNumber	(0008,0050)	N
IssuerOfAccessionNumberSequence/ LocalNameSpaceEntityID	(0008,0051)/ (0040,0031)	N
IssuerOfAccessionNumberSequence/ UniversalEntityID	(0008,0051)/ (0040,0032)	N
IssuerOfAccessionNumberSequence/ UniversalEntityIDType	(0008,0051)/ (0040,0033)	N
ReferringPhysicianName	(0008,0090)	N
ReferencedStudySequence/ ReferencedSOPClassUID	(0008,1110)/ (0008,1150)	N
ReferencedStudySequence/ ReferencedSOPInstanceUID	(0008,1110)/ (0008,1155)	N
PatientName	(0010,0010)	N
PatientID	(0010,0020)	N
IssuerOfPatientID	(0010,0021)	N
IssuerOfPatientIDQualifiersSequence/ UniversalEntityID	(0010,0024)/ (0040,0032)	N
IssuerOfPatientIDQualifiersSequence/ UniversalEntityIDType	(0010,0024)/ (0040,0033)	N
PatientBirthDate	(0010,0030)	N
PatientBirthTime	(0010,0032)	N
PatientSex	(0010,0040)	N
OtherPatientIDSequence/ PatientID	(0010,1002)/ (0010,0020)	N
OtherPatientIDSequence/ IssuerOfPatientID	(0010,1002)/ (0010,0021)	N
OtherPatientIDSequence/ TypeOfPatientID	(0010,1002)/ (0010,0022)	N
OtherPatientIDSequence/ IssuerOfPatientIDQualifiersSequence/ UniversalEntityID	(0010,1002)/ (0010,0024)/ (0040,0032)	N
OtherPatientIDSequence/ IssuerOfPatientIDQualifiersSequence/ UniversalEntityIDType	(0010,1002)/ (0010,0024)/ (0040,0033)	N
PatientBirthName	(0010,1005)	N
PatientMotherBirthName	(0010,1060)	N

Table 12: Supported C-FIND modality worklist keywords ▼

▲ Table 12: Supported C-FIND modality worklist keywords

Key	Tag	Used for matching
StudyInstanceUID	(0020,000D)	N
RequestedProcedureDescription	(0032,1060)	N
AdmissionID	(0038,0010)	N
IssuerOfAdmissionIDSequence/ LocalNamespaceEntityID	(0038,0014)/ (0040,0031)	N
IssuerOfAdmissionIDSequence/ UniversalEntityID	(0038,0014)/ (0040,0032)	N
IssuerOfAdmissionIDSequence/ UniversalEntityIDType	(0038,0014)/ (0040,0033)	N
ScheduledProcedureStepSequence/ ScheduledStationAETitle	(0040,0100)/ (0040,0001)	Y
ScheduledProcedureStepSequence/ ScheduledProcedureStepStartDate	(0040,0100)/ (0040,0002)	Y
ScheduledProcedureStepSequence/ ScheduledProcedureStepStartTime	(0040,0100)/ (0040,0003)	Y
ScheduledProcedureStepSequence/ ScheduledProcedureStepDescription	(0040,0100)/ (0040,0007)	N
ScheduledProcedureStepSequence/ ScheduledProcedureStepID	(0040,0100)/ (0040,0009)	N
ScheduledProcedureStepSequence/ ScheduledProtocolCodeSequence/ CodeValue	(0040,0100)/ (0040,0008)/ (0008,0100)	N
ScheduledProcedureStepSequence/ ScheduledProtocolCodeSequence/ CodeMeaning	(0040,0100)/ (0040,0008)/ (0008,0104)	N
ScheduledProcedureStepSequence/ Modality	(0040,0100)/ (0008,0060)	N
RequestedProcedureID	(0040,1001)	N

Manufacturer

ITH icoserve technology for healthcare GmbH
A Siemens Healthineers Company
Innrain 98
6020 Innsbruck
Austria
<https://www.ith-icoserve.com>

Distributed by

Siemens Healthineers AG
Siemensstr. 3
91301 Forchheim
Germany